



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115443

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/20/2021	16:20
2	CAL2	1	Water	NA	NA	21.2	7.01	07/20/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.03	07/20/2021	16:22
4	ICV	1	Water	NA	NA	21.2	6.99	07/20/2021	16:24
5	CCV1	1	Water	NA	NA	21.2	2.01	07/20/2021	16:25
6	M3098-01	1	Solid	20.02	20	23.8	7.35	07/20/2021	16:26
7	M3098-01DUP	1	Solid	20.03	20	23.9	7.36	07/20/2021	16:27
8	M3101-01	1	Solid	20.01	20	23.5	4.09	07/20/2021	16:30
9	CCV2	1	Water	NA	NA	21.3	12.02	07/20/2021	16:31

VB 115443

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3101 S WorkList ID : 150725 Department : Wet-Chemistry Date : 07-20-2021 16:12:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/20/2021	Solid	M3098-01	pH	Cool 4 deg C	PSEG03	L41	F0604S	07/20/2021	9045D
07/20/2021	Solid	M3101-01	pH	Cool 4 deg C	PSEG03	L41	33324	07/20/2021	9045D

Date/Time 07/20/21 16:15
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: JT (SM)

Date/Time 07/20/21 17:35
Raw Sample Received by: JT (SM)
Raw Sample Relinquished by: JO (WC)