



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115444

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/20/2021	16:20
2	CAL2	1	Water	NA	NA	21.2	7.01	07/20/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.03	07/20/2021	16:22
4	ICV	1	Water	NA	NA	21.2	6.99	07/20/2021	16:24
5	CCV1	1	Water	NA	NA	21.2	2.01	07/20/2021	16:25
6	M3102-01	1	Water	NA	NA	20.6	6.54	07/20/2021	16:35
7	M3102-03	1	Water	NA	NA	20.7	5.53	07/20/2021	16:36
8	M3102-04	1	Water	NA	NA	20.2	6.51	07/20/2021	16:37
9	M3102-05	1	Water	NA	NA	20.6	5.44	07/20/2021	16:40
10	M3102-06	1	Water	NA	NA	20.6	6.10	07/20/2021	16:42
11	M3103-01	1	Water	NA	NA	22.2	4.59	07/20/2021	16:45
12	M3104-01	1	Water	NA	NA	20.8	6.35	07/20/2021	16:46
13	M3104-03	1	Water	NA	NA	22.3	4.87	07/20/2021	16:49
14	M3104-03DUP	1	Water	NA	NA	22.4	4.88	07/20/2021	16:50
15	CCV2	1	Water	NA	NA	21.3	12.03	07/20/2021	17:00

WORKLIST(Hardcopy Internal Chain)

VB 115444

WorkList Name : PH M3101 W WorkList ID : 150724 Department : Wet-Chemistry Date : 07-20-2021 16:12:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/20/2021	Water	M3102-01	C pH	Cool 4 deg C	PSEG03	L41	33321	07/20/2021	9040C
07/20/2021	Water	M3102-03	pH	Cool 4 deg C	PSEG03	L41	33360	07/20/2021	9040C
07/20/2021	Water	M3102-04	pH	Cool 4 deg C	PSEG03	L41	33361	07/20/2021	9040C
07/20/2021	Water	M3102-05	pH	Cool 4 deg C	PSEG03	L41	33362	07/20/2021	9040C
07/20/2021	Water	M3102-06	pH	Cool 4 deg C	PSEG03	L41	33363	07/20/2021	9040C
07/20/2021	Water	M3103-01	A pH	Cool 4 deg C	PSEG03	L41	33307	07/20/2021	9040C
07/20/2021	Water	M3104-01	A pH	Cool 4 deg C	PSEG03	L41	33269	07/20/2021	9040C
07/20/2021	Water	M3104-03	A pH	Cool 4 deg C	PSEG03	L41	33270	07/20/2021	9040C

Date/Time 07/20/21 16:15
 Raw Sample Received by: JB (WC)
 Raw Sample Relinquished by: JT (SM)

Date/Time 07/20/21 17:35
 Raw Sample Received by: JT (SM)
 Raw Sample Relinquished by: JB (WC)