



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115454

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/21/2021	12:10
2	CAL2	1	Water	NA	NA	21.2	7.00	07/21/2021	12:11
3	CAL3	1	Water	NA	NA	21.2	10.03	07/21/2021	12:12
4	ICV	1	Water	NA	NA	21.3	6.99	07/21/2021	12:15
5	CCV1	1	Water	NA	NA	21.3	2.01	07/21/2021	12:16
6	M3116-01	1	Water	NA	NA	21.7	5.93	07/21/2021	12:19
7	M3116-01DUP	1	Water	NA	NA	21.8	5.94	07/21/2021	12:20
8	M3120-02	1	Water	NA	NA	20.3	5.03	07/21/2021	12:22
9	M3120-03	1	Water	NA	NA	20.4	5.49	07/21/2021	12:24
10	CCV2	1	Water	NA	NA	21.3	12.03	07/21/2021	12:25

1B 115454

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3119 WorkList ID : 150751 Department : Wet-Chemistry Date : 07-21-2021 11:54:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/20/2021	Water	M3116-01	pH	Cool 4 deg C	STER05	L11	BOTTLE-3	07/20/2021	9040C
07/19/2021	Water	M3120-02	pH	Cool 4 deg C	TETR06	L11	TT-RW6B-IDWGW-2021C	07/19/2021	9040C
07/19/2021	Water	M3120-03	pH	Cool 4 deg C	TETR06	L11	TT-RW6A-IDWGW-2021C	07/19/2021	9040C

Date/Time 07/21/21 12:00
 Raw Sample Received by: JD (Sul)
 Raw Sample Relinquished by: JD (Sul)

Date/Time 07/21/21 13:00
 Raw Sample Received by: JD (Sul)
 Raw Sample Relinquished by: JD (WC)