



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 07/22/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 07/21/2021 16:20 DATE & TIME OUT1: 07/22/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 07/22/2021 08:30 DATE & TIME OUT2: 07/22/2021 10:00Run Number: LB115461BalanceID: WC SC-3OvenID: Wc Oven-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB115461BL	LB115461BL	148.6917	148.6917	100	148.6918	148.6918	148.6918	0.0001	1
2	M3132-01	55-ST-32	152.7643	152.7643	100	152.9014	152.9014	152.9014	0.1371	1371
3	M3132-01DUP	55-ST-32DUP	152.9802	152.9802	100	153.1214	153.1214	153.1214	0.1412	1412

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

2B 115461

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M3132 WorkList ID : 150761 Department : Wet-Chemistry Date : 07-21-2021 15:36:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/28/2021	Water	M3132-01	TS	Cool 4 deg C	DIST01	L11	55-ST-32	07/21/2021	SM2540 B

Date/Time 07/21/21 15:40
Raw Sample Received by: JO (WC) CP SM
Raw Sample Relinquished by: CP SM

Date/Time 07/21/21 17:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JO