



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115465

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/21/2021	16:24
2	CAL2	1	Water	NA	NA	21.3	7.01	07/21/2021	16:25
3	CAL3	1	Water	NA	NA	21.3	10.02	07/21/2021	16:26
4	ICV	1	Water	NA	NA	21.3	6.99	07/21/2021	16:27
5	CCV1	1	Water	NA	NA	21.3	2.01	07/21/2021	16:30
6	M3128-02	1	Solid	20.02	20	23.3	7.75	07/21/2021	16:32
7	M3135-01	1	Solid	20.01	20	23.0	8.11	07/21/2021	16:35
8	M3135-04	1	Solid	20.02	20	22.6	7.89	07/21/2021	16:38
9	M3135-07	1	Solid	20.03	20	22.8	8.22	07/21/2021	16:40
10	M3135-07DUP	1	Solid	20.04	20	22.9	8.23	07/21/2021	16:41
11	CCV2	1	Water	NA	NA	21.4	12.03	07/21/2021	16:42

WORKLIST(Hardcopy Internal Chain)

LB 115465

WorkList Name : CORROSIONITY M3131 WorkList ID : 150763 Department : Wet-Chemistry Date : 07-21-2021 16:17:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/21/2021	Solid	M3128-02	Corrosivity	Cool 4 deg C	PSEG03	L51	ETGI-358	07/21/2021	9045D
07/21/2021	Solid	M3135-01	Corrosivity	Cool 4 deg C	PSEG03	L51	TP-A	07/21/2021	9045D
07/21/2021	Solid	M3135-04	Corrosivity	Cool 4 deg C	PSEG03	L51	TP-B	07/21/2021	9045D
07/21/2021	Solid	M3135-07	Corrosivity	Cool 4 deg C	PSEG03	L51	TP-L	07/21/2021	9045D

Date/Time 07/21/21 16:20
 Raw Sample Received by: JD (WC)
 Raw Sample Relinquished by: JD (CSM)

Date/Time 07/21/21 17:10
 Raw Sample Received by: JD (CSM)
 Raw Sample Relinquished by: JD (WC)