


7470A

LB115483
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38052



A= 0.0000e+000

B= 2.6232e-004

C= 1.4676e-002

Rho= 0.9995765

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.018	0.018	14	0.000	14					
0.2	0.200	0.202	0.002	716	0.0 %	716					
2.5	2.500	2.377	-0.123	9005	0.0 %	9005					
5.0	5.000	5.208	0.208	19798	0.0 %	19798					
7.5	7.500	7.398	-0.102	28147	0.0 %	28147					
10.0	10.000	9.996	-0.004	38052	0.0 %	38052					

100
1
5
4
1
0

LB115483 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jul 2021 14:24:22

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.0000	7470A	PPB	22 Jul 2021 14:27:03	S	Std
0.2 - 1	50.2	716	-	0.2000	7470A	PPB	22 Jul 2021 14:29:19	S	Std
2.5 - 1	52.5	9005	-	2.5000	7470A	PPB	22 Jul 2021 14:31:36	S	Std
5.0 - 1	55	19798	-	5.0000	7470A	PPB	22 Jul 2021 14:33:53	S	Std
7.5 - 1	57.5	28147	-	7.5000	7470A	PPB	22 Jul 2021 14:36:09	S	Std
10.0 - 1	59	38052	-	10.0000	7470A	PPB	22 Jul 2021 14:38:25	S	Std
ICV16 - 1	ICV16	14762	3.8870	-	7470A	PPB	22 Jul 2021 14:41:15	U	SMPL
ICB16 - 1	ICB16	-62	-0.0016	-	7470A	PPB	22 Jul 2021 14:43:30	U	SMPL
CCV09 - 1	CCV09	19582	5.1513	-	7470A	PPB	22 Jul 2021 14:45:44	U	SMPL
CCB09 - 1	CCB09	-88	-0.0084	-	7470A	PPB	22 Jul 2021 14:47:59	U	SMPL
CRA - 1	CRA	687	0.1949	-	7470A	PPB	22 Jul 2021 14:50:14	U	SMPL
HighStd - 1	HighStd	38388	10.0844	-	7470A	PPB	22 Jul 2021 14:52:29	U	SMPL
ChkStd - 1	ChkStd	26416	6.9440	-	7470A	PPB	22 Jul 2021 14:54:44	U	SMPL
PB137877BL - 1	PBW	-133	-0.0202	-	7470A	PPB	22 Jul 2021 14:57:01	U	SMPL
PB137877BS - 1	LCSW	13669	3.6003	-	7470A	PPB	22 Jul 2021 14:59:17	U	SMPL
M3114-07 - 1	20210594-AMENDED-COM-8-SPLP	-8	0.0126	-	7470A	PPB	22 Jul 2021 15:01:33	U	SMPL
M3114-07DUP - 1	20210594-AMENDED-COM-8-SPLPDUP	42	0.0257	-	7470A	PPB	22 Jul 2021 15:03:49	U	SMPL
M3114-07MS - 1	20210594-AMENDED-COM-8-SPLPMS	14522	3.8240	-	7470A	PPB	22 Jul 2021 15:06:06	U	SMPL
M3114-07MSD - 1	20210594-AMENDED-COM-8-SPLPMSD	14022	3.6929	-	7470A	PPB	22 Jul 2021 15:08:23	U	SMPL
M3118-05 - 1	20210588-SPLP-AMENDED-COM-2	-106	-0.0131	-	7470A	PPB	22 Jul 2021 15:10:39	U	SMPL
CCV10 - 1	CCV10	19524	5.1361	-	7470A	PPB	22 Jul 2021 15:12:54	U	SMPL
CCB10 - 1	CCB10	-110	-0.0142	-	7470A	PPB	22 Jul 2021 15:15:09	U	SMPL
M3119-02 - 1	TOP-SOIL-GRILLIO-1	402	0.1201	-	7470A	PPB	22 Jul 2021 15:17:24	U	SMPL
PB137878BL - 1	PBW	-30	0.0068	-	7470A	PPB	22 Jul 2021 15:19:39	U	SMPL
PB137878BS - 1	LCSW	12395	3.2661	-	7470A	PPB	22 Jul 2021 15:21:54	U	SMPL
M3128-02 - 1	ETGI-358	-17	0.0102	-	7470A	PPB	22 Jul 2021 15:24:09	U	SMPL
M3129-02 - 1	RO-VNJ-207	-8	0.0126	-	7470A	PPB	22 Jul 2021 15:26:24	U	SMPL
M3131-01 - 1	21194-21240	2	0.0152	-	7470A	PPB	22 Jul 2021 15:28:40	U	SMPL
M3135-02 - 1	TP-A	4	0.0157	-	7470A	PPB	22 Jul 2021 15:30:56	U	SMPL
M3135-05 - 1	TP-B	8	0.0168	-	7470A	PPB	22 Jul 2021 15:33:13	U	SMPL
M3135-08 - 1	TP-L	4	0.0157	-	7470A	PPB	22 Jul 2021 15:35:30	U	SMPL
M3135-08DUP - 1	TP-LDUP	24	0.0210	-	7470A	PPB	22 Jul 2021 15:37:46	U	SMPL
CCV11 - 1	CCV11	19587	5.1526	-	7470A	PPB	22 Jul 2021 15:40:01	U	SMPL
CCB11 - 1	CCB11	-84	-0.0074	-	7470A	PPB	22 Jul 2021 15:42:17	U	SMPL
M3135-08MS - 1	TP-LMS	13486	3.5523	-	7470A	PPB	22 Jul 2021 15:44:32	U	SMPL
M3135-08MSD - 1	TP-LMSD	14628	3.8518	-	7470A	PPB	22 Jul 2021 15:46:48	U	SMPL
PB137877TB - 1	PB137832TB	-112	-0.0147	-	7470A	PPB	22 Jul 2021 15:49:03	U	SMPL
PB137878TB - 1	PB137849TB	-11	0.0118	-	7470A	PPB	22 Jul 2021 15:51:19	U	SMPL
M3114-07LX5 - 1		22	0.0204	-	7470A	PPB	22 Jul 2021 15:53:35	U	SMPL
M3114-07A - 1		14617	3.8489	-	7470A	PPB	22 Jul 2021 15:55:51	U	SMPL
M3135-08LX5 - 1		-69	-0.0034	-	7470A	PPB	22 Jul 2021 15:58:07	U	SMPL
M3135-08A - 1		14890	3.9205	-	7470A	PPB	22 Jul 2021 16:00:24	U	SMPL
CCV12 - 1	CCV12	19567	5.1474	-	7470A	PPB	22 Jul 2021 16:02:40	U	SMPL
CCB12 - 1	CCB12	-103	-0.0123	-	7470A	PPB	22 Jul 2021 16:04:56	U	SMPL