



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115501

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/23/2021	13:15
2	CAL2	1	Water	NA	NA	21.2	7.01	07/23/2021	13:16
3	CAL3	1	Water	NA	NA	21.2	10.03	07/23/2021	13:17
4	ICV	1	Water	NA	NA	21.2	6.99	07/23/2021	13:19
5	CCV1	1	Water	NA	NA	21.3	2.01	07/23/2021	13:20
6	M3167-01	1	Solid	20.02	20	21.2	8.70	07/23/2021	13:22
7	M3167-01DUP	1	Solid	20.03	20	21.3	8.71	07/23/2021	13:23
8	CCV2	1	Water	NA	NA	21.3	12.02	07/23/2021	13:25

WB 115501

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3167 WorkList ID : 150839 Department : Wet-Chemistry Date : 07-23-2021 13:06:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/20/2021	Solid	M3167-01	pH	Cool 4 deg C	GEIC06	L11	TOP-SOIL-PILE-WHOLE	07/20/2021	9045D

Date/Time 07/23/21 13:10
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: Cef Sm

Date/Time 07/23/21 14:00
Raw Sample Received by: Cef
Raw Sample Relinquished by: JO Case 1