



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115509

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/23/2021	09:45
2	M3102-01		1	9	100.00	>212.0	>212.0	07/23/2021	10:10
3	M3102-01DUP		1	9	100.00	>212.0	>212.0	07/23/2021	10:35
4	M3102-07		1	8	100.00	>212.0	>212.0	07/23/2021	11:10
5	M3103-01		1	10	100.00	>212.0	>212.0	07/23/2021	11:40
6	M3104-01		1	9	100.00	>212.0	>212.0	07/23/2021	12:15
7	M3104-03		1	8	100.00	>212.0	>212.0	07/23/2021	12:40
8	M3132-01		1	10	100.00	>212.0	>212.0	07/23/2021	13:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 11 5309

WorkList Name : fp72321 WorkList ID : 150828 Department : Wet-Chemistry Date : 07-23-2021 09:15:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/27/2021	Water	M3102-01	Flash Point	Cool 4 deg C	PSEG03	L41	33321	07/20/2021	1010B
07/27/2021	Water	M3102-07	Flash Point	Cool 4 deg C	PSEG03	L41	60-63	07/20/2021	1010B
07/27/2021	Water	M3103-01	Flash Point	Cool 4 deg C	PSEG03	L41	33307	07/20/2021	1010B
07/27/2021	Water	M3104-01	Flash Point	Cool 4 deg C	PSEG03	L41	33269	07/20/2021	1010B
07/27/2021	Water	M3104-03	Flash Point	Cool 4 deg C	PSEG03	L41	33270	07/20/2021	1010B
08/03/2021	Water	M3132-01	Flash Point	Cool 4 deg C	DIST01	L11	55-ST-32	07/21/2021	1010B

Date/Time 07/23/21 09:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 07/23/21 13:10
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)