



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115515

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/23/2021	17:00
2	CAL2	1	Water	NA	NA	21.2	7.00	07/23/2021	17:01
3	CAL3	1	Water	NA	NA	21.2	10.02	07/23/2021	17:02
4	ICV	1	Water	NA	NA	21.3	6.99	07/23/2021	17:05
5	CCV1	1	Water	NA	NA	21.3	2.01	07/23/2021	17:06
6	M3171-01	1	Solid	20.03	20	22.9	7.56	07/23/2021	17:20
7	M3171-01DUP	1	Solid	20.04	20	22.8	7.57	07/23/2021	17:21
8	M3174-01	1	Solid	20.05	20	22.9	8.38	07/23/2021	17:24
9	CCV2	1	Water	NA	NA	21.3	12.02	07/23/2021	17:40

15515

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3171 WorkList ID : 150853 Department : Wet-Chemistry Date : 07-23-2021 16:34:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/23/2021	Solid	M3171-01	pH	Cool 4 deg C	TULL02	L53	FK-PS-01-078	07/23/2021	9045D
07/23/2021	Solid	M3174-01	pH	Cool 4 deg C	TULL02	L11	FK-EF-01-021-ABCD	07/23/2021	9045D

Date/Time 07/23/21 16:35
Raw Sample Received by: JP (W9)
Raw Sample Relinquished by: JP (W9)

Date/Time 07/23/21 17:15
Raw Sample Received by: JP (W9)
Raw Sample Relinquished by: JP (W9)