



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115516

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/23/2021	17:00
2	CAL2	1	Water	NA	NA	21.2	7.00	07/23/2021	17:01
3	CAL3	1	Water	NA	NA	21.2	10.03	07/23/2021	17:02
4	ICV	1	Water	NA	NA	21.3	6.99	07/23/2021	17:05
5	CCV1	1	Water	NA	NA	21.3	2.01	07/23/2021	17:06
6	M3148-03	1	Solid	20.02	20	22.0	10.06	07/23/2021	17:10
7	M3176-02	1	Solid	20.03	20	22.7	7.99	07/23/2021	17:12
8	M3176-02DUP	1	Solid	20.04	20	22.8	8.00	07/23/2021	17:13
9	CCV2	1	Water	NA	NA	21.3	12.02	07/23/2021	17:15

VB 115516

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M3168 WorkList ID : 150854 Department : Wet-Chemistry Date : 07-23-2021 16:35:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/22/2021	Solid	M3148-03	Corrosivity	Cool 4 deg C	FRAN01	L52	SI-2-EXCA-SOIL	07/22/2021	9045D
07/23/2021	Solid	M3176-02	Corrosivity	Cool 4 deg C	WALS01	L53	18-B7-TP-(9.5-12)	07/23/2021	9045D

Date/Time 07/23/21 1640
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JD CSM

Date/Time 07/23/21 1740
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JD CSM