

LB115526


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37761



A= 0.0000e+000

B= 2.6385e-004

C= 3.2955e-002

Rho= 0.9999720

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.027	0.027	-24	0.000	-24					
0.2	0.200	0.211	0.011	674	0.0 %	674					
2.5	2.500	2.443	-0.057	9136	0.0 %	9136					
5.0	5.000	5.000	-0.000	18824	0.0 %	18824					
7.5	7.500	7.523	0.023	28388	0.0 %	28388					
10.0	10.000	9.996	-0.004	37761	0.0 %	37761					

1
5
-2
0
0
0

LB115520 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Jul 2021 09:26:29				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-24	-	0.00007470A	PPB		26 Jul 2021 09:29:46	S Std
0.2 - 1	50.2	674	-	0.20007470A	PPB		26 Jul 2021 09:32:03	S Std
2.5 - 1	52.5	9136	-	2.50007470A	PPB		26 Jul 2021 09:34:20	S Std
5.0 - 1	55	18824	-	5.00007470A	PPB		26 Jul 2021 09:36:37	S Std
7.5 - 1	57.5	28388	-	7.50007470A	PPB		26 Jul 2021 09:38:53	S Std
10.0 - 1	510	37761	-	10.00007470A	PPB		26 Jul 2021 09:44:01	S Std
ICV19 - 1	ICV19	14813	3.9414	-7470A	PPB		26 Jul 2021 09:48:21	U SMPL
ICB19 - 1	ICB19	-154	-0.0077	-7470A	PPB		26 Jul 2021 09:50:37	U SMPL
CCV20 - 1	CCV20	18494	4.9126	-7470A	PPB		26 Jul 2021 09:52:51	U SMPL
CCB20 - 1	CCB20	-111	0.0037	-7470A	PPB		26 Jul 2021 09:55:06	U SMPL
CRA - 1	CRA	685	0.2137	-7470A	PPB		26 Jul 2021 09:57:22	U SMPL
HighStd - 1	HighStd	37240	9.8588	-7470A	PPB		26 Jul 2021 09:59:36	U SMPL
ChkStd - 1	ChkStd	26830	7.1121	-7470A	PPB		26 Jul 2021 10:01:51	U SMPL
PB137919BL - 1	PBW	-161	-0.0095	-7470A	PPB		26 Jul 2021 10:04:08	U SMPL
PB137919BS - 1	LCSW	12845	3.4221	-7470A	PPB		26 Jul 2021 10:06:24	U SMPL
M3148-03 - 1	SI-2-EXCA-SOIL	-93	0.0084	-7470A	PPB		26 Jul 2021 10:08:41	U SMPL
M3150-02 - 1	TP-B	43	0.0443	-7470A	PPB		26 Jul 2021 10:10:58	U SMPL
M3150-05 - 1	TP-C	50	0.0461	-7470A	PPB		26 Jul 2021 10:13:15	U SMPL
M3152-02 - 1	VAULT-A	54	0.0472	-7470A	PPB		26 Jul 2021 10:15:32	U SMPL
M3152-05 - 1	VAULT-B	44	0.0446	-7470A	PPB		26 Jul 2021 10:17:47	U SMPL
CCV21 - 1	CCV21	18844	5.0050	-7470A	PPB		26 Jul 2021 10:20:02	U SMPL
CCB21 - 1	CCB21	-74	0.0134	-7470A	PPB		26 Jul 2021 10:22:17	U SMPL
M3152-08 - 1	COLL-HOLE-A	8	0.0351	-7470A	PPB		26 Jul 2021 10:24:33	U SMPL
M3152-11 - 1	COLL-HOLE-B	8	0.0351	-7470A	PPB		26 Jul 2021 10:26:48	U SMPL
M3152-11DUP - 1	COLL-HOLE-BDUP	21	0.0385	-7470A	PPB		26 Jul 2021 10:29:03	U SMPL
M3152-11MS - 1	COLL-HOLE-BMS	15279	4.0643	-7470A	PPB		26 Jul 2021 10:31:19	U SMPL
M3152-11MSD - 1	COLL-HOLE-BMSD	16510	4.3891	-7470A	PPB		26 Jul 2021 10:33:34	U SMPL
PB137919TB - 1	PB1137849TB	-65	0.0158	-7470A	PPB		26 Jul 2021 10:35:51	U SMPL
M3152-11LX5 - 1		-6	0.0314	-7470A	PPB		26 Jul 2021 10:38:08	U SMPL
M3152-11A - 1		14005	3.7282	-7470A	PPB		26 Jul 2021 10:40:24	U SMPL
CCV22 - 1	CCV22	18652	4.9543	-7470A	PPB		26 Jul 2021 10:42:41	U SMPL
CCB22 - 1	CCB22	-117	0.0021	-7470A	PPB		26 Jul 2021 10:44:57	U SMPL