



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115530

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	07/26/2021	12:15
2	CAL2	1	Water	NA	NA	21.5	7.00	07/26/2021	12:16
3	CAL3	1	Water	NA	NA	21.5	10.04	07/26/2021	12:17
4	ICV	1	Water	NA	NA	21.6	6.99	07/26/2021	12:19
5	CCV1	1	Water	NA	NA	21.4	2.01	07/26/2021	12:20
6	M3186-01	1	Solid	20.03	20	25.6	7.22	07/26/2021	12:22
7	M3186-03	1	Solid	20.04	20	25.3	7.19	07/26/2021	13:25
8	M3186-05	1	Solid	20.02	20	25.5	7.69	07/26/2021	13:29
9	M3186-05DUP	1	Solid	20.01	20	25.6	7.70	07/26/2021	13:30
10	CCV2	1	Water	NA	NA	21.6	12.02	07/26/2021	13:32

SP 115530

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3186 WorkList ID : 150898 Department : Wet-Chemistry Date : 07-26-2021 12:08:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/26/2021	Solid	M3186-01	pH	Cool 4 deg C	DASK01	L61	AREA-1-(0-5)	07/26/2021	9045D
07/26/2021	Solid	M3186-03	pH	Cool 4 deg C	DASK01	L61	AREA-2-(0-7)	07/26/2021	9045D
07/26/2021	Solid	M3186-05	pH	Cool 4 deg C	DASK01	L61	AREA-3-(0-9)	07/26/2021	9045D

Date/Time 07/26/21 12:10
Raw Sample Received by: JG (WC)
Raw Sample Relinquished by: JD CSM

Date/Time 07/26/21 12:20
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JG (WC)