



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115533

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	07/26/2021	13:10
2	CAL2	1	Water	NA	NA	21.5	7.00	07/26/2021	13:11
3	CAL3	1	Water	NA	NA	21.5	10.03	07/26/2021	13:12
4	ICV	1	Water	NA	NA	21.6	7.01	07/26/2021	13:16
5	CCV1	1	Water	NA	NA	21.6	2.01	07/26/2021	13:19
6	M3196-01	1	Water	NA	NA	25.1	5.90	07/26/2021	13:20
7	M3196-01DUP	1	Water	NA	NA	25.2	5.91	07/26/2021	13:21
8	M3197-01	1	Water	NA	NA	24.7	5.81	07/26/2021	13:24
9	CCV2	1	Water	NA	NA	21.6	12.02	07/26/2021	13:25

VB 115533

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3195 w WorkList ID : 150901 Department : Wet-Chemistry Date : 07-26-2021 12:58:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/15/2021	Water	M3196-01	pH	Cool 4 deg C	STER14	L61	MINI-CELL	07/15/2021	9040C
07/15/2021	Water	M3197-01	pH	Cool 4 deg C	STER14	L61	JS10000	07/15/2021	9040C

Date/Time 07/26/21 13:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: CP SM

Date/Time 07/26/21 14:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WC)