

LB115534


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

56232



A= 0.0000e+000

B= 1.7736e-004

C= 4.5907e-002

Rho= 0.9999702

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.050	0.050	22	0.000	22					
0.2	0.200	0.162	-0.038	653	0.0 %	653					
2.5	2.500	2.502	0.002	13847	0.0 %	13847					
5.0	5.000	4.980	-0.020	27819	0.0 %	27819					
7.5	7.500	7.488	-0.012	41961	0.0 %	41961					
10.0	10.000	10.019	0.019	56232	0.0 %	56232					

40)

-19

0

0

0

0

LB115534 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 26 Jul 2021 14:15:06					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	22	-	0.0000	7471B	PPB	26 Jul 2021 14:17:25	S	Std
0.2 - 1	50.2	653	-	0.2000	7471B	PPB	26 Jul 2021 14:19:41	S	Std
2.5 - 1	52.5	13847	-	2.5000	7471B	PPB	26 Jul 2021 14:21:59	S	Std
5.0 - 1	55	27819	-	5.0000	7471B	PPB	26 Jul 2021 14:24:15	S	Std
7.5 - 1	57.5	41961	-	7.5000	7471B	PPB	26 Jul 2021 14:26:31	S	Std
10.0 - 1	510	56232	-	10.0000	7471B	PPB	26 Jul 2021 14:28:48	S	Std
ICV21 - 1	ICV21	22116	3.9683	-	7471B	PPB	26 Jul 2021 14:31:38	U	SMPL
ICB21 - 1	ICB21	-83	0.0312	-	7471B	PPB	26 Jul 2021 14:33:52	U	SMPL
CCV27 - 1	CCV27	27728	4.9636	-	7471B	PPB	26 Jul 2021 14:36:08	U	SMPL
CCB27 - 1	CCB27	-71	0.0333	-	7471B	PPB	26 Jul 2021 14:38:23	U	SMPL
CRA - 1	CRA	638	0.1591	-	7471B	PPB	26 Jul 2021 14:40:39	U	SMPL
HighStd - 1	HighStd	55870	9.9548	-	7471B	PPB	26 Jul 2021 14:42:54	U	SMPL
ChkStd - 1	ChkStd	39181	6.9949	-	7471B	PPB	26 Jul 2021 14:45:09	U	SMPL
PB137976BL - 1	PBS	-234	0.0044	-	7471B	PPB	26 Jul 2021 14:47:26	U	SMPL
PB137976BS - 1	LCSS	22219	3.9866	-	7471B	PPB	26 Jul 2021 14:52:18	U	SMPL
M3165-01 - 1	COMP-1	2756	0.5347	-	7471B	PPB	26 Jul 2021 14:59:31	U	SMPL
M3165-01DUP - 1	COMP-1DUP	2617	0.5100	-	7471B	PPB	26 Jul 2021 15:01:46	U	SMPL
M3165-01MS - 1	COMP-1MS	23396	4.1953	-	7471B	PPB	26 Jul 2021 15:04:03	U	SMPL
M3165-01MSD - 1	COMP-1MSD	22844	4.0974	-	7471B	PPB	26 Jul 2021 15:08:02	U	SMPL
M3165-02 - 1	COMP-2	1581	0.3263	-	7471B	PPB	26 Jul 2021 15:10:16	U	SMPL
CCV28 - 1	CCV28	27374	4.9009	-	7471B	PPB	26 Jul 2021 15:12:32	U	SMPL
CCB28 - 1	CCB28	-87	0.0305	-	7471B	PPB	26 Jul 2021 15:14:47	U	SMPL
M3165-03 - 1	COMP-3	806	0.1889	-	7471B	PPB	26 Jul 2021 15:17:02	U	SMPL
M3170-01 - 1	SU-02-072321	1125	0.2454	-	7471B	PPB	26 Jul 2021 15:19:17	U	SMPL
M3171-01 - 1	FK-PS-01-078	2259	0.4466	-	7471B	PPB	26 Jul 2021 15:21:32	U	SMPL
M3174-01 - 1	FK-EF-01-021-ABCD	1100	0.2410	-	7471B	PPB	26 Jul 2021 15:23:47	U	SMPL
M3175-01 - 1	FK-BPM-03-025-ABCD	2014	0.4031	-	7471B	PPB	26 Jul 2021 15:26:07	U	SMPL
M3176-01 - 1	18-B7-TP-(9.5-12)	9079	1.6561	-	7471B	PPB	26 Jul 2021 15:28:23	U	SMPL
M3165-01LX5 - 1		486	0.1321	-	7471B	PPB	26 Jul 2021 15:30:40	U	SMPL
M3165-01A - 1		24538	4.3979	-	7471B	PPB	26 Jul 2021 15:32:57	U	SMPL
CCV29 - 1	CCV29	27686	4.9562	-	7471B	PPB	26 Jul 2021 15:35:15	U	SMPL
CCB29 - 1	CCB29	-81	0.0315	-	7471B	PPB	26 Jul 2021 15:37:35	U	SMPL