



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/27/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:47
In Date: 07/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:20
Out Date: 07/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115538

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3183-01	VNJ-201-RBR-200036	1	1.15	9.78	9.36	95.1	
M3183-03	VNJ-201-RBR-200036-EPH-2	2	1.15	9.78	9.36	95.1	
M3184-01	20210419-5	3	1.12	9.66	6.97	68.5	sediment sample
M3184-02	20210482-68	4	1.13	9.98	6.69	62.8	sediment sample
M3184-03	20210494-80	5	1.14	6.66	3.36	40.2	sediment sample
M3184-04	20210530-COM-3	6	1.16	9.57	6.17	59.6	sediment sample
M3184-05	20210561-COM-34	7	1.11	9.87	6.12	57.2	sediment sample
M3184-06	20210564-COM-37	8	1.15	9.97	5.13	45.1	sediment sample
M3186-01	AREA-1- (0-5)	9	1.16	9.56	8.98	93.1	
M3186-02	AREA-A- (5)	10	1.16	9.96	9.36	93.2	
M3186-03	AREA-2- (0-7)	11	1.18	9.71	9.36	95.9	
M3186-04	AREA-2- (7)	12	1.11	9.75	9.38	95.7	
M3186-05	AREA-3- (0-9)	13	1.16	9.97	9.6	95.8	
M3186-06	AREA-3- (9)	14	1.1	9.94	9.22	91.9	
M3193-01	01-A-01-B-01-C	15	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-02	02-A-02-B-02-C	16	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-03	03-A-03-B-03-C	17	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-04	04-A-04-B-04-C	18	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-05	05-A-05-B-05-C	19	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-06	06-A-06-B-06-C	20	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-07	07-A-07-B-07-C	21	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-08	08-A-07-B-07-C	22	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-09	09-A-09-B-09-C	23	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-10	10-A-10-B-10-C	24	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-11	11-A-11-B-11-C	25	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-12	12-A-12-B-12-C	26	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-13	13-A-13-B-13-C	27	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3193-14	14-A-14-B-14-C	28	1.00	2.00	2.00	100.0	CAULKING SAMPLE



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/27/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:47
In Date: 07/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:20
Out Date: 07/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115538

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3194-01	JC-03-072621	29	1.13	9.72	9.03	92.0	
M3194-02	JC-03-072621-EPH-2	30	1.18	9.91	9.48	95.1	
M3199-01	HD-02-072621	31	1.16	9.67	8.6	87.4	
M3199-02	HD-02-072621-EPH-2	32	1.13	9.61	8.7	89.3	
M3200-01	GS-1	33	1.19	9.87	8.92	89.1	
M3200-02	GS-1A	34	1.16	9.53	8.73	90.4	
M3200-03	GS-2	35	1.19	9.71	9.00	91.7	
M3200-04	GS-2A	36	1.15	9.98	9.06	89.6	
M3200-05	GS-3	37	1.15	9.97	9.17	90.9	
M3200-06	GS-3A	38	1.13	9.52	8.8	91.4	
M3200-06D	GS-3ADUP	39	1.16	9.54	8.83	91.5	
M3201-01	P001-B47C07-01	40	1.00	2.00	2.00	100.0	CONCRETE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$