

LB115539


SFAM01.1

INSTRUMENT 1D1.CV1

Linear

 μ Abs.:

43050



A= 0.0000e+000

B= 2.2990e-004 *slope*C= 6.6015e-002 *y-intercept*

Rho= 0.9995858

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/b)</i>
0.00	0.000	0.083	0.083	72	0.000	72					<i>2</i>
0.05	0.050					573					<i>-7</i>
0.20	0.200	0.187	-0.013	528	0.0 %	528					<i>-1</i>
2.50	2.500	2.478	-0.022	10493	0.0 %	10493					<i>-4</i>
5.00	5.000	4.823	-0.177	20690	0.0 %	20690					<i>2</i>
7.50	7.500	7.666	0.166	33056	0.0 %	33056					<i>0</i>
10.0	10.000	9.963	-0.037	43050	0.0 %	43050					

LB115539

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	72	-		0 SFAM01.1	PPB	7/26/2021 16:10	Std
	0.2 S01	528	-		0.2 SFAM01.1	PPB	7/26/2021 16:12	Std
	2.5 S02	10493	-		2.5 SFAM01.1	PPB	7/26/2021 16:14	Std
	5 S03	20690	-		5 SFAM01.1	PPB	7/26/2021 16:17	Std
	7.5 S04	33056	-		7.5 SFAM01.1	PPB	7/26/2021 16:23	Std
	10 S05	43050	-		10 SFAM01.1	PPB	7/26/2021 16:31	Std
ICV051	ICV051	17618	4.1164 -		SFAM01.1	PPB	7/26/2021 16:33	SMPL
ICB051	ICB051	-75	0.0488 -		SFAM01.1	PPB	7/26/2021 16:36	SMPL
CCV030	CCV030	21439	4.9949 -		SFAM01.1	PPB	7/26/2021 16:38	SMPL
CCB030	CCB030	-91	0.0451 -		SFAM01.1	PPB	7/26/2021 16:40	SMPL
PB137978BL	PBW978	3	0.0667 -		SFAM01.1	PPB	7/26/2021 16:45	SMPL
PB137978TB	LEB978	12	0.0688 -		SFAM01.1	PPB	7/26/2021 16:47	SMPL
M3154-01	MC0AA8	310	0.1373 -		SFAM01.1	PPB	7/26/2021 16:49	SMPL
M3154-02	MC0AB0	17	0.0699 -		SFAM01.1	PPB	7/26/2021 16:57	SMPL
M3154-03	MC0AB0D	16	0.0697 -		SFAM01.1	PPB	7/26/2021 16:59	SMPL
M3154-04	MC0AB0S	4465	1.0925 -		SFAM01.1	PPB	7/26/2021 17:01	SMPL
CCV031	CCV031	20686	4.8218 -		SFAM01.1	PPB	7/26/2021 17:03	SMPL
CCB031	CCB031	-112	0.0403 -		SFAM01.1	PPB	7/26/2021 17:06	SMPL