



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115540

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/26/2021	16:35
2	M3177-01		1	9	100.00	>212.0	>212.0	07/26/2021	17:05
3	M3177-01DUP		1	9	100.00	>212.0	>212.0	07/26/2021	17:40
4	M3195-01		1	8	100.00	>212.0	>212.0	07/26/2021	18:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 115340

WorkList Name : FP72621 WorkList ID : 150909 Department : Wet-Chemistry Date : 07-26-2021 16:05:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/26/2021	Water	M3177-01	B Flash Point	Cool 4 deg C	WALS01	L53	WATER-SAMPLE	07/23/2021	1010B
08/09/2021	Water	M3195-01	F Flash Point	Cool 4 deg C	PSEG03	L61	SUMPWATER-TRE-1232	07/26/2021	1010B

Date/Time 07/26/21 16:10
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 07/26/21 18:10
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)