


LB115559

SFAM01.1

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.5767e-004 slope

C= -5.8517e-002 y-intercept

Rho= 0.9998823

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.045	-0.045	52	0.000	52					-
0.05	0.050					573					-9
0.20	0.200	0.183	-0.017	938	0.0 %	938					1
2.50	2.500	2.530	0.030	10044	0.0 %	10044					2
5.00	5.000	5.111	0.111	20063	0.0 %	20063					-1
7.50	7.500	7.443	-0.057	29114	0.0 %	29114					0
10.0	10.000	9.978	-0.022	38951	0.0 %	38951					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	52	-		0 SFAM01.1	PPB	7/27/2021 14:00	Std
	0.2 S01	938	-		0.2 SFAM01.1	PPB	7/27/2021 14:02	Std
	2.5 S02	10044	-		2.5 SFAM01.1	PPB	7/27/2021 14:07	Std
	5 S03	20063	-		5 SFAM01.1	PPB	7/27/2021 14:09	Std
	7.5 S04	29114	-		7.5 SFAM01.1	PPB	7/27/2021 14:14	Std
	10 S05	38951	-		10 SFAM01.1	PPB	7/27/2021 14:21	Std
ICV052	ICV052	16579	4.2134	-	SFAM01.1	PPB	7/27/2021 14:24	SMPL
ICB052	ICB052	25	-0.0521	-	SFAM01.1	PPB	7/27/2021 14:26	SMPL
CCV032	CCV032	19309	4.9168	-	SFAM01.1	PPB	7/27/2021 14:28	SMPL
CCB032	CCB032	17	-0.0541	-	SFAM01.1	PPB	7/27/2021 14:31	SMPL
PB137996BL	PBW996	9	-0.0562	-	SFAM01.1	PPB	7/27/2021 14:33	SMPL
M3112-17	MCOAX7	-24	-0.0647	-	SFAM01.1	PPB	7/27/2021 14:35	SMPL
PB137997BL	PBW997	23	-0.0526	-	SFAM01.1	PPB	7/27/2021 14:37	SMPL
M3113-01	MCOAD2	-13	-0.0619	-	SFAM01.1	PPB	7/27/2021 14:40	SMPL
M3113-02	MCOAD8	1320	0.2816	-	SFAM01.1	PPB	7/27/2021 14:42	SMPL
M3113-03	MCOAD9	1615	0.3576	-	SFAM01.1	PPB	7/27/2021 14:44	SMPL
M3113-04	MCOAE0	1045	0.2107	-	SFAM01.1	PPB	7/27/2021 14:46	SMPL
M3113-05	MCOAE1	198	-0.0075	-	SFAM01.1	PPB	7/27/2021 14:49	SMPL
M3113-06	MCOAD4	174	-0.0137	-	SFAM01.1	PPB	7/27/2021 14:51	SMPL
M3113-07	MCOAD7	1652	0.3672	-	SFAM01.1	PPB	7/27/2021 14:53	SMPL
M3113-08	MCOAY3	-28	-0.0657	-	SFAM01.1	PPB	7/27/2021 14:55	SMPL
M3113-09	MCOAY3D	-33	-0.067	-	SFAM01.1	PPB	7/27/2021 14:58	SMPL
M3113-10	MCOAY3S	4427	1.0822	-	SFAM01.1	PPB	7/27/2021 15:00	SMPL
M3113-11	MCOAY4	-29	-0.066	-	SFAM01.1	PPB	7/27/2021 15:02	SMPL
M3113-12	MCOAY2	-20	-0.0637	-	SFAM01.1	PPB	7/27/2021 15:04	SMPL
PB137998BL	PBW998	22	-0.0528	-	SFAM01.1	PPB	7/27/2021 15:07	SMPL
M3115-01	MCOAE4	8	-0.0565	-	SFAM01.1	PPB	7/27/2021 15:09	SMPL
M3115-02	MCOAF0	-34	-0.0673	-	SFAM01.1	PPB	7/27/2021 15:11	SMPL
M3115-03	MCOAF1	-21	-0.0639	-	SFAM01.1	PPB	7/27/2021 15:14	SMPL
M3115-04	MCOAF2	-15	-0.0624	-	SFAM01.1	PPB	7/27/2021 15:16	SMPL
CCV033	CCV033	19050	4.8501	-	SFAM01.1	PPB	7/27/2021 15:18	SMPL
CCB033	CCB033	26	-0.0518	-	SFAM01.1	PPB	7/27/2021 15:20	SMPL
M3115-05	MCOAF3	-34	-0.0673	-	SFAM01.1	PPB	7/27/2021 15:23	SMPL
M3115-06	MCOAE6	-13	-0.0619	-	SFAM01.1	PPB	7/27/2021 15:25	SMPL
M3115-07	MCOAE7	5	-0.0572	-	SFAM01.1	PPB	7/27/2021 15:27	SMPL
M3115-08	MCOAE8	31	-0.0505	-	SFAM01.1	PPB	7/27/2021 15:29	SMPL
M3115-09	MCOAE8D	32	-0.0503	-	SFAM01.1	PPB	7/27/2021 15:32	SMPL
M3115-10	MCOAE8S	4166	1.0149	-	SFAM01.1	PPB	7/27/2021 15:34	SMPL
M3115-11	MCOAE9	-39	-0.0686	-	SFAM01.1	PPB	7/27/2021 15:36	SMPL
M3115-12	MCOAA5	-6	-0.0601	-	SFAM01.1	PPB	7/27/2021 15:38	SMPL
CCV034	CCV034	19037	4.8468	-	SFAM01.1	PPB	7/27/2021 15:41	SMPL
CCB034	CCB034	24	-0.0523	-	SFAM01.1	PPB	7/27/2021 15:43	SMPL