



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115560

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	07/26/2021	16:25
2	CAL2	1	Water	NA	NA	21.6	7.00	07/26/2021	16:26
3	CAL3	1	Water	NA	NA	21.6	10.03	07/26/2021	16:27
4	ICV	1	Water	NA	NA	21.8	7.02	07/26/2021	16:30
5	CCV1	1	Water	NA	NA	21.7	2.01	07/26/2021	16:32
6	M3195-01	1	Water	NA	NA	23.7	7.60	07/26/2021	16:35
7	M3195-01DUP	1	Water	NA	NA	23.8	7.61	07/26/2021	16:36
8	CCV2	1	Water	NA	NA	21.7	12.02	07/26/2021	16:38

WP 115560

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3195

WorkList ID : 150945

Department : Wet-Chemistry

Date : 07-26-21 16:17:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/26/2021	Water	M3195-01	C pH	Cool 4 deg C	PSEG03	L61	SUMPWATER-TRE-1232	07/26/2021	9040C

Date/Time 07/26/21 16:20

Raw Sample Received by: JP CWC

Raw Sample Relinquished by: JP CWC

Date/Time 07/26/21 17:30

Raw Sample Received by: JP CWC

Raw Sample Relinquished by: JP CWC