

LB 115565


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51496



A= 0.00

B= 1.91

C= 3.15

Rho= 0.99

Accept=Acce

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	39	0.000	39				
0.2	0.200	0.193	-0.007	1006	0.0 %	1006				
2.5	2.500	2.624	0.124	13668	0.0 %	13668				
5.0	5.000	4.710	-0.290	24533	0.0 %	24533				
7.5	7.500	7.780	0.280	40526	0.0 %	40526				
10.0	10.000	9.885	-0.115	51496	0.0 %	51496				

Method: 7471B

Operator: Admin

Date of Analysis: 27 Jul 2021 15:56:14

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units		
0.0 - 1	50	39	-	0.00007471B	PPB		27 Jul 2021	15:58
0.2 - 1	50.2	1006	-	0.20007471B	PPB		27 Jul 2021	16:00
2.5 - 1	52.5	13668	-	2.50007471B	PPB		27 Jul 2021	16:08
5.0 - 1	55	24533	-	5.00007471B	PPB		27 Jul 2021	16:07
7.5 - 1	57.5	40526	-	7.50007471B	PPB		27 Jul 2021	16:10
10.0 - 1	510	51496	-	10.00007471B	PPB		27 Jul 2021	16:12
ICV23 - 1	ICV23	21755	4.1764	-7471B	PPB		27 Jul 2021	16:15
ICB23 - 1	ICB23	-62	-0.0116	-7471B	PPB		27 Jul 2021	16:17
CCV35 - 1	CCV35	24618	4.7260	-7471B	PPB		27 Jul 2021	16:19
CCB35 - 1	CCB35	-62	-0.0116	-7471B	PPB		27 Jul 2021	16:21
CRA - 1	CRA	991	0.1905	-7471B	PPB		27 Jul 2021	16:24
HighStd - 1	HighStd	51283	9.8446	-7471B	PPB		27 Jul 2021	16:26
ChkStd - 1	ChkStd	37360	7.1719	-7471B	PPB		27 Jul 2021	16:28
PB138007BL - 1	PBS	-177	-0.0337	-7471B	PPB		27 Jul 2021	16:30
PB138007BS - 1	LCSS	19181	3.6823	-7471B	PPB		27 Jul 2021	16:33
M3183-01 - 1	VNJ-201-RBR-200036	5017	0.9634	-7471B	PPB		27 Jul 2021	16:35
M3184-04 - 1	20210530-COM-3	23594	4.5294	-7471B	PPB		27 Jul 2021	16:37
M3184-05 - 1	20210561-COM-34	17450	3.3500	-7471B	PPB		27 Jul 2021	16:38
M3184-06 - 1	20210564-COM-37	38581	7.4063	-7471B	PPB		27 Jul 2021	16:42
M3186-01 - 1	AREA-1-(0-5)	31820	6.1085	-7471B	PPB		27 Jul 2021	16:44
CCV36 - 1	CCV36	24456	4.6949	-7471B	PPB		27 Jul 2021	16:46
CCB36 - 1	CCB36	-65	-0.0122	-7471B	PPB		27 Jul 2021	16:48
M3186-03 - 1	AREA-2-(0-7)	42322	8.1244	-7471B	PPB		27 Jul 2021	16:51
M3186-05 - 1	AREA-3-(0-9)	15329	2.9429	-7471B	PPB		27 Jul 2021	16:53
M3194-01 - 1	JC-03-072621	3880	0.7451	-7471B	PPB		27 Jul 2021	16:55
M3199-01 - 1	HD-02-072621	644	0.1239	-7471B	PPB		27 Jul 2021	16:58
M3200-01 - 1	GS-1	2077	0.3990	-7471B	PPB		27 Jul 2021	17:00
M3200-03 - 1	GS-2	1681	0.3230	-7471B	PPB		27 Jul 2021	17:02
M3200-05 - 1	GS-3	1461	0.2808	-7471B	PPB		27 Jul 2021	17:04
M3200-05DUP - 1	GS-3DUP	1321	0.2539	-7471B	PPB		27 Jul 2021	17:07
M3200-05MS - 1	GS-3MS	20694	3.9727	-7471B	PPB		27 Jul 2021	17:09
M3200-05MSD - 1	GS-3MSD	20665	3.9672	-7471B	PPB		27 Jul 2021	17:11
CCV37 - 1	CCV37	24614	4.7252	-7471B	PPB		27 Jul 2021	17:13
CCB37 - 1	CCB37	-55	-0.0102	-7471B	PPB		27 Jul 2021	17:16
M3200-05LX5 - 1		295	0.0569	-7471B	PPB		27 Jul 2021	17:18
M3200-05A - 1		22375	4.2954	-7471B	PPB		27 Jul 2021	17:20
CCV38 - 1	CCV38	24584	4.7194	-7471B	PPB		27 Jul 2021	17:22
CCB38 - 1	CCB38	-70	-0.0131	-7471B	PPB		27 Jul 2021	17:25