



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115568

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/27/2021	14:15
2	CAL2	1	Water	NA	NA	21.3	7.00	07/27/2021	14:16
3	CAL3	1	Water	NA	NA	21.4	10.03	07/27/2021	14:19
4	ICV	1	Water	NA	NA	21.5	6.98	07/27/2021	14:20
5	CCV1	1	Water	NA	NA	21.6	2.01	07/27/2021	14:22
6	M3183-02	1	Solid	20.03	20	25.2	9.41	07/27/2021	14:25
7	M3183-02DUP	1	Solid	20.04	20	25.3	9.42	07/27/2021	14:26
8	CCV2	1	Water	NA	NA	21.6	12.03	07/27/2021	14:30

VB 115568

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M3183 WorkList ID : 150946 Department : Wet-Chemistry Date : 07-27-2021 14:06:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/26/2021	Solid	M3183-02	Corrosivity	Cool 4 deg C	PSEG03	L61	VNJ-201-RBR-200036	07/26/2021	9045D

Date/Time 07/27/21 14:10
Raw Sample Received by: JR WJC
Raw Sample Relinquished by: JR WJC

Date/Time 07/27/21 15:30
Raw Sample Received by: JR WJC
Raw Sample Relinquished by: JR WJC