


LB115575

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4427e-004

C= 1.4216e-002

Rho= 0.9999893

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.028 | 0.028  | 57    | 0.000      | 57    |       |       |       |       |    |
| 0.2    | 0.200  | 0.183 | -0.017 | 691   | 0.0 %      | 691   |       |       |       |       | -9 |
| 2.5    | 2.500  | 2.483 | -0.017 | 10108 | 0.0 %      | 10108 |       |       |       |       | -1 |
| 5.0    | 5.000  | 4.993 | -0.007 | 20384 | 0.0 %      | 20384 |       |       |       |       | 0  |
| 7.5    | 7.500  | 7.518 | 0.018  | 30718 | 0.0 %      | 30718 |       |       |       |       | 0  |
| 10.0   | 10.000 | 9.994 | -0.006 | 40856 | 0.0 %      | 40856 |       |       |       |       | 8  |

## LB115575 INSTRUMENT ID : CV1

| Method: 7470A   |                       | Operator: Admin |         | Date of Analysis: 28 Jul 2021 10:54:00 |        |       |                      |      |      |
|-----------------|-----------------------|-----------------|---------|----------------------------------------|--------|-------|----------------------|------|------|
| Sample ID       | Extended ID           | μ Abs.          | Conc.   | Std Conc                               | Method | Units | Date                 | Type | Type |
| 0.0 - 1         | 50                    | 57              | -       | 0.00007470A                            |        | PPB   | 28 Jul 2021 11:05:09 | S    | Std  |
| 0.2 - 1         | 50.2                  | 691             | -       | 0.20007470A                            |        | PPB   | 28 Jul 2021 11:07:26 | S    | Std  |
| 2.5 - 1         | 52.5                  | 10108           | -       | 2.50007470A                            |        | PPB   | 28 Jul 2021 11:09:42 | S    | Std  |
| 5.0 - 1         | 55                    | 20384           | -       | 5.00007470A                            |        | PPB   | 28 Jul 2021 11:14:33 | S    | Std  |
| 7.5 - 1         | 57.5                  | 30718           | -       | 7.50007470A                            |        | PPB   | 28 Jul 2021 11:19:50 | S    | Std  |
| 10.0 - 1        | 57.5                  | 40856           | -       | 10.00007470A                           |        | PPB   | 28 Jul 2021 11:27:20 | S    | Std  |
| ICV25 - 1       | ICV25                 | 15002           | 3.6788  | -7470A                                 |        | PPB   | 28 Jul 2021 11:30:01 | U    | SMPL |
| ICB25 - 1       | ICB25                 | -62             | -0.0009 | -7470A                                 |        | PPB   | 28 Jul 2021 11:34:45 | U    | SMPL |
| CCV39 - 1       | CCV39                 | 20251           | 4.9610  | -7470A                                 |        | PPB   | 28 Jul 2021 11:37:00 | U    | SMPL |
| CCB39 - 1       | CCB39                 | -113            | -0.0134 | -7470A                                 |        | PPB   | 28 Jul 2021 11:39:15 | U    | SMPL |
| CRA - 1         | CRA                   | 665             | 0.1767  | -7470A                                 |        | PPB   | 28 Jul 2021 11:41:30 | U    | SMPL |
| HighStd - 1     | HighStd               | 41078           | 10.0485 | -7470A                                 |        | PPB   | 28 Jul 2021 11:43:45 | U    | SMPL |
| ChkStd - 1      | ChkStd                | 29596           | 7.2437  | -7470A                                 |        | PPB   | 28 Jul 2021 11:46:00 | U    | SMPL |
| PB138004BL - 1  | PBW                   | -6              | 0.0128  | -7470A                                 |        | PPB   | 28 Jul 2021 11:50:47 | U    | SMPL |
| PB138004BS - 1  | LCSW                  | 15242           | 3.7374  | -7470A                                 |        | PPB   | 28 Jul 2021 11:53:03 | U    | SMPL |
| M3176-02 - 1    | 18-B7-TP-(9.5-12)     | 222             | 0.0684  | -7470A                                 |        | PPB   | 28 Jul 2021 11:55:19 | U    | SMPL |
| M3183-02 - 1    | VNJ-201-RBR-200036    | 302             | 0.0880  | -7470A                                 |        | PPB   | 28 Jul 2021 11:57:36 | U    | SMPL |
| M3183-02DUP - 1 | VNJ-201-RBR-200036DUP | 315             | 0.0912  | -7470A                                 |        | PPB   | 28 Jul 2021 11:59:53 | U    | SMPL |
| M3183-02MS - 1  | VNJ-201-RBR-200036MS  | 15411           | 3.7787  | -7470A                                 |        | PPB   | 28 Jul 2021 12:02:10 | U    | SMPL |
| M3183-02MSD - 1 | VNJ-201-RBR-200036MSD | 14566           | 3.5723  | -7470A                                 |        | PPB   | 28 Jul 2021 12:04:25 | U    | SMPL |
| CCV40 - 1       | CCV40                 | 20555           | 5.0353  | -7470A                                 |        | PPB   | 28 Jul 2021 12:06:40 | U    | SMPL |
| CCB40 - 1       | CCB40                 | -69             | -0.0026 | -7470A                                 |        | PPB   | 28 Jul 2021 12:08:55 | U    | SMPL |
| PB138005BL - 1  | PBW                   | 9               | 0.0164  | -7470A                                 |        | PPB   | 28 Jul 2021 12:11:11 | U    | SMPL |
| PB138005BS - 1  | LCSW                  | 14358           | 3.5215  | -7470A                                 |        | PPB   | 28 Jul 2021 12:13:26 | U    | SMPL |
| M3195-01 - 1    | SUMPWATER-TRE-1232    | 238             | 0.0724  | -7470A                                 |        | PPB   | 28 Jul 2021 12:15:41 | U    | SMPL |
| M3195-01DUP - 1 | SUMPWATER-TRE-1232DUP | 322             | 0.0929  | -7470A                                 |        | PPB   | 28 Jul 2021 12:20:41 | U    | SMPL |
| M3195-01MS - 1  | SUMPWATER-TRE-1232MS  | 15526           | 3.8068  | -7470A                                 |        | PPB   | 28 Jul 2021 12:22:56 | U    | SMPL |
| M3195-01MSD - 1 | SUMPWATER-TRE-1232MSD | 15352           | 3.7643  | -7470A                                 |        | PPB   | 28 Jul 2021 12:25:12 | U    | SMPL |
| PB138013BL - 1  | PBW                   | 6               | 0.0157  | -7470A                                 |        | PPB   | 28 Jul 2021 12:30:27 | U    | SMPL |
| PB138013BS - 1  | LCSW                  | 14571           | 3.5735  | -7470A                                 |        | PPB   | 28 Jul 2021 12:32:43 | U    | SMPL |
| M3156-11 - 1    | MW-1                  | 209             | 0.0653  | -7470A                                 |        | PPB   | 28 Jul 2021 12:35:00 | U    | SMPL |
| M3156-11DUP - 1 | MW-1DUP               | 252             | 0.0758  | -7470A                                 |        | PPB   | 28 Jul 2021 12:37:15 | U    | SMPL |
| CCV41 - 1       | CCV41                 | 20420           | 5.0023  | -7470A                                 |        | PPB   | 28 Jul 2021 12:39:31 | U    | SMPL |
| CCB41 - 1       | CCB41                 | -84             | -0.0063 | -7470A                                 |        | PPB   | 28 Jul 2021 12:41:46 | U    | SMPL |
| M3156-11MS - 1  | MW-1MS                | 14709           | 3.6072  | -7470A                                 |        | PPB   | 28 Jul 2021 12:44:02 | U    | SMPL |
| M3156-11MSD - 1 | MW-1MSD               | 14994           | 3.6768  | -7470A                                 |        | PPB   | 28 Jul 2021 12:46:17 | U    | SMPL |
| M3156-12 - 1    | MW-1                  | 135             | 0.0472  | -7470A                                 |        | PPB   | 28 Jul 2021 12:48:32 | U    | SMPL |
| PB138004TB - 1  | PB137958TB            | 305             | 0.0887  | -7470A                                 |        | PPB   | 28 Jul 2021 12:50:47 | U    | SMPL |
| PB138005TB - 1  | PB137990TB            | 268             | 0.0797  | -7470A                                 |        | PPB   | 28 Jul 2021 12:53:03 | U    | SMPL |
| CCV42 - 1       | CCV42                 | 20187           | 4.9454  | -7470A                                 |        | PPB   | 28 Jul 2021 12:55:19 | U    | SMPL |
| CCB42 - 1       | CCB42                 | -61             | -0.0007 | -7470A                                 |        | PPB   | 28 Jul 2021 12:57:35 | U    | SMPL |
| M3183-02LX5 - 1 |                       | 283             | 0.0833  | -7470A                                 |        | PPB   | 28 Jul 2021 12:59:52 | U    | SMPL |
| M3183-02A - 1   |                       | 15837           | 3.8828  | -7470A                                 |        | PPB   | 28 Jul 2021 13:02:09 | U    | SMPL |
| M3195-01LX5 - 1 |                       | 43              | 0.0247  | -7470A                                 |        | PPB   | 28 Jul 2021 13:04:24 | U    | SMPL |
| M3195-01A - 1   |                       | 15790           | 3.8713  | -7470A                                 |        | PPB   | 28 Jul 2021 13:06:40 | U    | SMPL |
| M3156-11LX5 - 1 |                       | 24              | 0.0201  | -7470A                                 |        | PPB   | 28 Jul 2021 13:08:55 | U    | SMPL |
| M3156-11A - 1   |                       | 15991           | 3.9204  | -7470A                                 |        | PPB   | 28 Jul 2021 13:11:10 | U    | SMPL |
| CCV43 - 1       | CCV43                 | 20434           | 5.0057  | -7470A                                 |        | PPB   | 28 Jul 2021 13:13:26 | U    | SMPL |
| CCB43 - 1       | CCB43                 | -97             | -0.0095 | -7470A                                 |        | PPB   | 28 Jul 2021 13:15:41 | U    | SMPL |