



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 07/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 07/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115578

| Lab ID | Client SampleID | Dish # | Dish Wt (g) (A) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|--------------------------|----------------------------|---------|----------|
| M3180-01 | BG2L7 | 1 | 1.14 | 9.92 | 9.04 | 90.0 | |
| M3180-02 | BG2L7D | 2 | 1.14 | 9.92 | 9.04 | 90.0 | |
| M3180-03 | BG2L7S | 3 | 1.14 | 9.92 | 9.04 | 90.0 | |
| M3180-04 | BG290 | 4 | 1.16 | 9.6 | 7.39 | 73.8 | |
| M3180-05 | BG2A8 | 5 | 1.1 | 9.89 | 8.21 | 80.9 | |
| M3180-06 | BG2A9 | 6 | 1.16 | 9.92 | 8.66 | 85.6 | |
| M3180-07 | BG2C3 | 7 | 1.13 | 9.82 | 8.42 | 83.9 | |
| M3180-08 | BG2C6 | 8 | 1.16 | 9.91 | 7.81 | 76.0 | |
| M3180-09 | BG2D8 | 9 | 1.13 | 9.91 | 8.42 | 83.0 | |
| M3180-10 | BG2E1 | 10 | 1.16 | 9.62 | 7.77 | 78.1 | |
| M3180-11 | BG2J9 | 11 | 1.11 | 9.58 | 8.41 | 86.2 | |
| M3180-12 | BG2K2 | 12 | 1.16 | 9.98 | 7.9 | 76.4 | |
| M3180-13 | BG2K5 | 13 | 1.18 | 9.6 | 7.86 | 79.3 | |
| M3180-14 | BG2K8 | 14 | 1.13 | 9.62 | 7.38 | 73.6 | |
| M3180-15 | BG2L1 | 15 | 1.19 | 9.82 | 8.11 | 80.2 | |
| M3180-16 | BG2M0 | 16 | 1.15 | 9.58 | 7.95 | 80.7 | |
| M3180-17 | BG2Y4 | 17 | 1.17 | 9.79 | 8.32 | 82.9 | |
| M3180-18 | BG2Y5 | 18 | 1.12 | 9.74 | 6.92 | 67.3 | |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$