

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : *AM*

Instrument ID : Konelab

7/28/2021 13:05

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.095	0.0	0.098	
ICB1	0.001	0.0	0.004	
CCV1	1.073	0.0	0.096	
CCB1	0.001	0.0	0.004	
PB137963BL	0.000	0.0	0.004	
PB137963BS	1.104	0.0	0.099	
M2940-12	0.028	0.0	0.006	
M3110-01	0.048	0.0	0.008	
M3110-01DUP	0.046	0.0	0.007	
M3110-01MS	1.096	0.0	0.098	
M3110-01MSD	1.100	0.0	0.099	
M3132-01	0.050	0.0	0.008	
CCV2	1.079	0.0	0.097	
CCB2	0.000	0.0	0.004	

N	14
Mean	0.480
SD	0.5495
CV%	114.42

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 28 12:09:37 2021

Wed Jul 28 13:03:59 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0007	mg/l	7/28/2021 12:09:37	
0.05PPM	A	Phenolics-	P	0.0442	mg/l	7/28/2021 12:09:38	
0.1PPM	A	Phenolics-	P	0.1052	mg/l	7/28/2021 12:09:39	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	7/28/2021 12:09:40	
0.50PPM	A	Phenolics-	P	0.5022	mg/l	7/28/2021 12:09:41	
1.0PPM	A	Phenolics-	P	0.9932	mg/l	7/28/2021 12:09:42	
2.0PPM	A	Phenolics-	P	2.0025	mg/l	7/28/2021 12:09:43	
ICV1	S	Phenolics-	P	1.0954	mg/l	7/28/2021 12:58:05	
ICB1	S	Phenolics-	P	0.001	mg/l	7/28/2021 12:58:07	
CCV1	S	Phenolics-	P	1.0728	mg/l	7/28/2021 12:58:09	
CCB1	S	Phenolics-	P	0.0012	mg/l	7/28/2021 12:58:10	
PB137963BL	S	Phenolics-	P	0.0004	mg/l	7/28/2021 12:58:12	
PB137963BS	S	Phenolics-	P	1.1041	mg/l	7/28/2021 12:58:13	
M2940-12	S	Phenolics-	P	0.0283	mg/l	7/28/2021 12:58:14	
M3110-01	S	Phenolics-	P	0.0484	mg/l	7/28/2021 12:58:15	
M3110-01DUP	S	Phenolics-	P	0.0462	mg/l	7/28/2021 13:03:54	
M3110-01MS	S	Phenolics-	P	1.0965	mg/l	7/28/2021 13:03:55	
M3110-01MSD	S	Phenolics-	P	1.0998	mg/l	7/28/2021 13:03:56	
M3132-01	S	Phenolics-	P	0.0499	mg/l	7/28/2021 13:03:57	
CCV2	S	Phenolics-	P	1.0789	mg/l	7/28/2021 13:03:58	
CCB2	S	Phenolics-	P	0.0004	mg/l	7/28/2021 13:03:59	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AK

Instrument ID : Konelab

7/28/2021 12:25

Test Phenolics-


Accepted 7/28/2021 12:25

Factor 11.56

Bias 0.003

Coeff. of det. 0.999961

Errors



	Calibrator	Response	Calc. con.	Conc.	Relative Errors
1	0.00PPM	0.004	0.0007	0.0000	-11.6
2	Phenol-2	0.007	0.0442	0.0500	5.2
3	Phenol-2	0.013	0.1052	0.1000	0.8
4	Phenol-2	0.025	0.2519	0.2500	0.4
5	Phenol-2	0.047	0.5022	0.5000	-0.7
6	Phenol-2	0.089	0.9932	1.0000	0.1
7	Phenol-2	0.177	2.0025	2.0000	

AK
7/28/21