



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115585

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/28/2021	16:30
2	CAL2	1	Water	NA	NA	21.3	7.00	07/28/2021	16:32
3	CAL3	1	Water	NA	NA	21.4	10.02	07/28/2021	16:33
4	ICV	1	Water	NA	NA	21.5	7.00	07/28/2021	16:35
5	CCV1	1	Water	NA	NA	21.4	2.01	07/28/2021	16:36
6	M3221-06	1	Water	NA	NA	19.5	5.32	07/28/2021	16:38
7	M3221-06DUP	1	Water	NA	NA	19.5	5.33	07/28/2021	16:39
8	M3221-07	1	Water	NA	NA	17.7	5.25	07/28/2021	16:41
9	M3221-08	1	Water	NA	NA	18.5	5.40	07/28/2021	16:43
10	M3221-09	1	Water	NA	NA	20.3	5.34	07/28/2021	16:45
11	CCV2	1	Water	NA	NA	21.5	12.02	07/28/2021	16:47

WORKLIST(Hardcopy Internal Chain)

LB115385

WorkList Name : PH72821 WorkList ID : 150988 Department : Wet-Chemistry Date : 07-28-2021 16:19:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/28/2021	Water	M3221-06	pH	Cool 4 deg C	NYUL01	D11	TH-4-30-DI-1	07/28/2021	SM 4500-H B
07/28/2021	Water	M3221-07	pH	Cool 4 deg C	NYUL01	D11	TH-4-30-DI-2	07/28/2021	SM 4500-H B
07/28/2021	Water	M3221-08	pH	Cool 4 deg C	NYUL01	D11	TH-4-30-DI-3	07/28/2021	SM 4500-H B
07/28/2021	Water	M3221-09	pH	Cool 4 deg C	NYUL01	D11	TH-4-04-DI-4	07/28/2021	SM 4500-H B

Date/Time 07/28/21 16:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JR (WC)

Date/Time 07/28/21 16:50
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: KS (WC)