


LB 115605

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1247e-004 *slope*C= -1.5331e-002 *+ intercept*

Rho= 0.9998533

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.055	0.055	333	0.000	333					
0.05	0.050					573					
0.20	0.200	0.210	0.010	1059	0.0 %	1059					
2.50	2.500	2.500	-0.000	11837	0.0 %	11837					
5.00	5.000	4.866	-0.134	22975	0.0 %	22975					
7.50	7.500	7.519	0.019	35460	0.0 %	35460					
10.0	10.000	10.050	0.050	47375	0.0 %	47375					

115605

LB115605
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	333	-		0 SFAM01.1	PPB	7/29/2021 13:28	Std
	0.2 S01	1059	-		0.2 SFAM01.1	PPB	7/29/2021 13:31	Std
	2.5 S02	11837	-		2.5 SFAM01.1	PPB	7/29/2021 13:33	Std
	5 S03	22975	-		5 SFAM01.1	PPB	7/29/2021 13:35	Std
	7.5 S04	35460	-		7.5 SFAM01.1	PPB	7/29/2021 13:37	Std
	10 S05	47375	-		10 SFAM01.1	PPB	7/29/2021 13:40	Std
ICV053	ICV053	19095	4.0417 -		SFAM01.1	PPB	7/29/2021 13:43	SMPL
ICB053	ICB053	234	0.0344 -		SFAM01.1	PPB	7/29/2021 13:45	SMPL
CCV050	CCV050	23014	4.8744 -		SFAM01.1	PPB	7/29/2021 13:47	SMPL
CCB050	CCB050	252	0.0382 -		SFAM01.1	PPB	7/29/2021 13:49	SMPL
PB138057BL	PBS057	53	-0.0041 -		SFAM01.1	PPB	7/29/2021 13:52	SMPL
M3112-01	MCOAA9	2449	0.505 -		SFAM01.1	PPB	7/29/2021 13:54	SMPL
M3112-03	MCOAQ8	2391	0.4927 -		SFAM01.1	PPB	7/29/2021 13:56	SMPL
M3112-04	MCOAS7	3609	0.7515 -		SFAM01.1	PPB	7/29/2021 13:58	SMPL
M3112-05	MCOAT3	1322	0.2656 -		SFAM01.1	PPB	7/29/2021 14:01	SMPL
M3112-06	MCOAQ3	2002	0.41 -		SFAM01.1	PPB	7/29/2021 14:03	SMPL
M3112-07	MCOAQ4	2145	0.4404 -		SFAM01.1	PPB	7/29/2021 14:05	SMPL
M3112-08	MCOAQ5	1578	0.3199 -		SFAM01.1	PPB	7/29/2021 14:07	SMPL
M3112-09	MCOAQ6	1027	0.2029 -		SFAM01.1	PPB	7/29/2021 14:10	SMPL
M3112-10	MCOAQ7	1718	0.3497 -		SFAM01.1	PPB	7/29/2021 14:12	SMPL
M3112-11	MCOAS8	2587	0.5343 -		SFAM01.1	PPB	7/29/2021 14:14	SMPL
M3112-12	MCOAS9	3218	0.6684 -		SFAM01.1	PPB	7/29/2021 14:16	SMPL
M3112-13	MCOAT0	1296	0.26 -		SFAM01.1	PPB	7/29/2021 14:19	SMPL
M3112-14	MCOAT1	782	0.1508 -		SFAM01.1	PPB	7/29/2021 14:21	SMPL
M3112-15	MCOAT2	6381	1.3404 -		SFAM01.1	PPB	7/29/2021 14:23	SMPL
M3112-16	MCOAX2	2816	0.583 -		SFAM01.1	PPB	7/29/2021 14:25	SMPL
M3112-18	MCOAA8	806	0.1559 -		SFAM01.1	PPB	7/29/2021 14:28	SMPL
M3112-19	MCOAB0	754	0.1449 -		SFAM01.1	PPB	7/29/2021 14:30	SMPL
M3112-20	MCOAB0D	626	0.1177 -		SFAM01.1	PPB	7/29/2021 14:32	SMPL
M3112-21	MCOAB0S	12486	2.6375 -		SFAM01.1	PPB	7/29/2021 14:35	SMPL
CCV051	CCV051	22967	4.8644 -		SFAM01.1	PPB	7/29/2021 14:37	SMPL
CCB051	CCB051	240	0.0357 -		SFAM01.1	PPB	7/29/2021 14:39	SMPL
M3112-22	MCOAQ9	1866	0.3811 -		SFAM01.1	PPB	7/29/2021 14:41	SMPL
PB138058BL	PBS058	44	-0.006 -		SFAM01.1	PPB	7/29/2021 14:44	SMPL
M3127-01	DBK69	823	0.1595 -		SFAM01.1	PPB	7/29/2021 14:46	SMPL
M3127-02	DBK69D	807	0.1561 -		SFAM01.1	PPB	7/29/2021 14:48	SMPL
M3127-03	DBK69S	12930	2.7319 -		SFAM01.1	PPB	7/29/2021 14:50	SMPL
M3127-04	DBK70	124	0.011 -		SFAM01.1	PPB	7/29/2021 14:53	SMPL
M3127-05	DBK71	589	0.1098 -		SFAM01.1	PPB	7/29/2021 14:55	SMPL
PB138059BL	PBS259	67	-0.0011 -		SFAM01.1	PPB	7/29/2021 14:57	SMPL
M3155-01	MCOAR0	3121	0.6478 -		SFAM01.1	PPB	7/29/2021 14:59	SMPL
M3155-02	MCOAR1	1825	0.3724 -		SFAM01.1	PPB	7/29/2021 15:02	SMPL
M3155-03	MCOAS5	1422	0.2868 -		SFAM01.1	PPB	7/29/2021 15:04	SMPL
M3155-04	MCOAT4	1636	0.3323 -		SFAM01.1	PPB	7/29/2021 15:06	SMPL
M3155-05	MCOAT5	2636	0.5447 -		SFAM01.1	PPB	7/29/2021 15:08	SMPL
M3155-06	MCOAT5D	2376	0.4895 -		SFAM01.1	PPB	7/29/2021 15:11	SMPL

LB115605
INSTRUMENT ID : CV1

M3155-07	MC0AT5S	13906	2.9392 -	SFAM01.1	PPB	7/29/2021 15:16 SMPL
M3155-08	MC0AT6	986	0.1942 -	SFAM01.1	PPB	7/29/2021 15:18 SMPL
M3155-09	MC0AK4	2216	0.4555 -	SFAM01.1	PPB	7/29/2021 15:21 SMPL
M3155-10	MC0AK5	790	0.1525 -	SFAM01.1	PPB	7/29/2021 15:23 SMPL
M3155-11	MC0AK6	2627	0.5428 -	SFAM01.1	PPB	7/29/2021 15:25 SMPL
M3155-12	MC0AM4	4462	0.9327 -	SFAM01.1	PPB	7/29/2021 15:27 SMPL
CCV052	CCV052	22814	4.8319 -	SFAM01.1	PPB	7/29/2021 15:30 SMPL
CCB052	CCB052	248	0.0374 -	SFAM01.1	PPB	7/29/2021 15:32 SMPL
M3155-13	MC0AM8	3595	0.7485 -	SFAM01.1	PPB	7/29/2021 15:34 SMPL
M3155-14	MC0AK7	3323	0.6907 -	SFAM01.1	PPB	7/29/2021 15:36 SMPL
M3155-15	MC0AK9	2510	0.518 -	SFAM01.1	PPB	7/29/2021 15:39 SMPL
M3155-16	MC0AN0	6482	1.3619 -	SFAM01.1	PPB	7/29/2021 15:41 SMPL
M3155-17	MC0AN4	2367	0.4876 -	SFAM01.1	PPB	7/29/2021 15:43 SMPL
CCV053	CCV053	22918	4.854 -	SFAM01.1	PPB	7/29/2021 15:45 SMPL
CCB053	CCB053	249	0.0376 -	SFAM01.1	PPB	7/29/2021 15:48 SMPL