


LB115617

7471B

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.001	-0.001	150	0.000	150					100
0.2	0.200	0.186	-0.014	1130	0.0 %	1130					-7
2.5	2.500	2.426	-0.074	12922	0.0 %	12922					-3
5.0	5.000	5.234	0.234	27701	0.0 %	27701					5
7.5	7.500	7.346	-0.154	38814	0.0 %	38814					-2
10.0	10.000	10.009	0.009	52829	0.0 %	52829					0

LB115617 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 30 Jul 2021 09:10:34					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	150	-	0.0000	7471B	PPB	30 Jul 2021 09:14:24	S	Std
0.2 - 1	50.2	1130	-	0.2000	7471B	PPB	30 Jul 2021 09:16:42	S	Std
2.5 - 1	52.5	12922	-	2.5000	7471B	PPB	30 Jul 2021 09:18:59	S	Std
5.0 - 1	55	27701	-	5.0000	7471B	PPB	30 Jul 2021 09:21:16	S	Std
7.5 - 1	57.5	38814	-	7.5000	7471B	PPB	30 Jul 2021 09:23:32	S	Std
10.0 - 1	510	52829	-	10.0000	7471B	PPB	30 Jul 2021 09:28:33	S	Std
ICV29 - 1	ICV29	20406	3.8482	-	7471B	PPB	30 Jul 2021 09:31:11	U	SMPL
ICB29 - 1	ICB29	47	-0.0202	-	7471B	PPB	30 Jul 2021 09:33:26	U	SMPL
CCV54 - 1	CCV54	27512	5.1984	-	7471B	PPB	30 Jul 2021 09:35:42	U	SMPL
CCB54 - 1	CCB54	57	-0.0183	-	7471B	PPB	30 Jul 2021 09:37:56	U	SMPL
CRA - 1	CRA	1130	0.1856	-	7471B	PPB	30 Jul 2021 09:40:12	U	SMPL
HighStd - 1	HighStd	53302	10.0987	-	7471B	PPB	30 Jul 2021 09:42:27	U	SMPL
ChkStd - 1	ChkStd	36569	6.9193	-	7471B	PPB	30 Jul 2021 09:44:43	U	SMPL
PB138074BL - 1	PBS	21	-0.0251	-	7471B	PPB	30 Jul 2021 09:47:00	U	SMPL
PB138074BS - 1	LCSS	18014	3.3937	-	7471B	PPB	30 Jul 2021 09:49:17	U	SMPL
M3206-02 - 1	20210581-COM-4	43274	8.1933	-	7471B	PPB	30 Jul 2021 09:51:34	U	SMPL
M3206-03 - 1	20210590-AMENDED-COM-4	31110	5.8820	-	7471B	PPB	30 Jul 2021 09:53:50	U	SMPL
M3206-04 - 1	20210533-COM-6	15938	2.9992	-	7471B	PPB	30 Jul 2021 09:56:08	U	SMPL
M3210-01 - 1	SB-1(1.0-1.5)	1999	0.3507	-	7471B	PPB	30 Jul 2021 09:58:26	U	SMPL
M3210-01DUP - 1	SB-1(1.0-1.5)DUP	2283	0.4047	-	7471B	PPB	30 Jul 2021 10:00:41	U	SMPL
CCV55 - 1	CCV55	27508	5.1976	-	7471B	PPB	30 Jul 2021 10:02:56	U	SMPL
CCB55 - 1	CCB55	81	-0.0137	-	7471B	PPB	30 Jul 2021 10:05:12	U	SMPL
M3210-01MS - 1	SB-1(1.0-1.5)MS	22376	4.2225	-	7471B	PPB	30 Jul 2021 10:07:28	U	SMPL
M3210-01MSD - 1	SB-1(1.0-1.5)MSD	23249	4.3884	-	7471B	PPB	30 Jul 2021 10:09:43	U	SMPL
M3210-02 - 1	SB-2(1.0-1.5)	25206	4.7602	-	7471B	PPB	30 Jul 2021 10:11:59	U	SMPL
M3210-03 - 1	SB-3(1.0-1.5)	1979	0.3469	-	7471B	PPB	30 Jul 2021 10:14:16	U	SMPL
M3210-04 - 1	SB-4(1.0-2.0)	3594	0.6538	-	7471B	PPB	30 Jul 2021 10:16:32	U	SMPL
M3210-05 - 1	SB-5(1.0-5.0)	8885	1.6591	-	7471B	PPB	30 Jul 2021 10:18:48	U	SMPL
M3216-05 - 1	20210551-COM-24	4895	0.9010	-	7471B	PPB	30 Jul 2021 10:21:04	U	SMPL
M3216-06 - 1	20210567-COM-40	39590	7.4933	-	7471B	PPB	30 Jul 2021 10:23:21	U	SMPL
M3216-07 - 1	20210538-COM-11	50236	9.5161	-	7471B	PPB	30 Jul 2021 10:25:38	U	SMPL
M3216-08 - 1	20210584-COM-7	42658	8.0762	-	7471B	PPB	30 Jul 2021 10:27:55	U	SMPL
CCV56 - 1	CCV56	28089	5.3080	-	7471B	PPB	30 Jul 2021 10:30:11	U	SMPL
CCB56 - 1	CCB56	84	-0.0132	-	7471B	PPB	30 Jul 2021 10:32:28	U	SMPL
M3216-09 - 1	20210593-AMENDED-COM-7	35778	6.7690	-	7471B	PPB	30 Jul 2021 10:34:44	U	SMPL
M3216-10 - 1	20210565-COM-38	38988	7.3789	-	7471B	PPB	30 Jul 2021 10:36:59	U	SMPL
M3217-02 - 1	280	8103	1.5105	-	7471B	PPB	30 Jul 2021 10:39:16	U	SMPL
M3222-01 - 1	MH-458-COMP	402	0.0472	-	7471B	PPB	30 Jul 2021 10:41:33	U	SMPL
M3223-01 - 1	BACKFILL-ASHCROFT-REAR-09	234	0.0153	-	7471B	PPB	30 Jul 2021 10:43:49	U	SMPL
M3223-03 - 1	BACKFILL-ASHCROFT-REAR-10	224	0.0134	-	7471B	PPB	30 Jul 2021 10:46:05	U	SMPL
M3223-05 - 1	BACKFILL-ASHCROFT-REAR-11	222	0.0130	-	7471B	PPB	30 Jul 2021 10:48:20	U	SMPL
M3230-01 - 1	ETGI-357	3771	0.6874	-	7471B	PPB	30 Jul 2021 10:50:37	U	SMPL
M3210-01LX5 - 1		582	0.0815	-	7471B	PPB	30 Jul 2021 10:52:54	U	SMPL
M3210-01A - 1		23135	4.3667	-	7471B	PPB	30 Jul 2021 10:55:10	U	SMPL
CCV57 - 1	CCV57	27919	5.2757	-	7471B	PPB	30 Jul 2021 10:57:25	U	SMPL
CCB57 - 1	CCB57	155	0.0003	-	7471B	PPB	30 Jul 2021 10:59:41	U	SMPL