



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115618

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	07/30/2021	09:00
2	CAL2	1	Water	NA	NA	22.1	7.00	07/30/2021	09:01
3	CAL3	1	Water	NA	NA	22.1	10.03	07/30/2021	09:02
4	ICV	1	Water	NA	NA	22.2	6.99	07/30/2021	09:05
5	CCV1	1	Water	NA	NA	22.1	2.01	07/30/2021	09:06
6	M3229-04	1	Water	NA	NA	26.1	7.03	07/30/2021	09:10
7	M3229-04DUP	1	Water	NA	NA	26.2	7.04	07/30/2021	09:11
8	M3229-05	1	Water	NA	NA	25.1	7.11	07/30/2021	09:15
9	CCV2	1	Water	NA	NA	22.2	12.02	07/30/2021	09:17

28 115618

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3229 w WorkList ID : 151034 Department : Wet-Chemistry Date : 07-30-2021 08:48:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/29/2021	Water	M3229-04	pH	Cool 4 deg C	PSEG03	D12	WATER-A	07/29/2021	9040C
07/29/2021	Water	M3229-05	pH	Cool 4 deg C	PSEG03	D12	WATER-B	07/29/2021	9040C

Date/Time 07/30/21 08:50
Raw Sample Received by: JR (WC)
Raw Sample Relinquished by: JR (WC)

Date/Time 07/30/21 10:00
Raw Sample Received by: JR (SM)
Raw Sample Relinquished by: JR (WC)