



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115624

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	07/30/2021	13:35
2	CAL2	1	Water	NA	NA	20.3	7.01	07/30/2021	13:36
3	CAL3	1	Water	NA	NA	20.3	10.03	07/30/2021	13:37
4	ICV	1	Water	NA	NA	20.4	7.02	07/30/2021	13:39
5	CCV1	1	Water	NA	NA	20.4	2.01	07/30/2021	13:40
6	M3242-01	1	Water	NA	NA	20.1	7.00	07/30/2021	13:42
7	M3242-01DUP	1	Water	NA	NA	20.2	7.01	07/30/2021	13:43
8	CCV2	1	Water	NA	NA	20.4	12.02	07/30/2021	13:45

2B 115624

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH M3241 WorkList ID: 151051 Department: Wet-Chemistry Date: 07-30-2021 13:27:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/30/2021	Water	M3242-01	pH	Cool 4 deg C	PSEG03	D21	257392	07/30/2021	9040C

Date/Time 07/30/21 13:30
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 07/30/21 14:30
Raw Sample Received by: JP (SM)
Raw Sample Relinquished by: JP (WC)