



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115634

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	07/30/2021	13:45
2	M3229-06		1	9	100.00	>212.0	>212.0	07/30/2021	14:20
3	M3229-06DUP		1	9	100.00	>212.0	>212.0	07/30/2021	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115634

WorkList Name : FP73021 WorkList ID : 151049 Department : Wet-Chemistry Date : 07-30-2021 12:11:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/03/2021	Water	M3229-06	Flash Point	Cool 4 deg C	PSEG03	D12	WATER-A-B	07/29/2021	1010B

Date/Time 07/30/21 13:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CFA-1006

Date/Time 07/30/21 14:55
 Raw Sample Received by: RJ CFA-1006
 Raw Sample Relinquished by: KS (WC)