

L3115655


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

41232



A= 0.0000e+000

B= 2.4121e-004 *slope*

C= -7.7173e-003 *y-intercept*

Rho= 0.9999348

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.003	-0.003	18	0.000	18					
0.05	0.050					573					
0.20	0.200	0.185	-0.015	797	0.0 %	797					
2.50	2.500	2.497	-0.003	10383	0.0 %	10383					
5.00	5.000	5.004	0.004	20777	0.0 %	20777					
7.50	7.500	7.581	0.081	31460	0.0 %	31460					
10.0	10.000	9.938	-0.062	41232	0.0 %	41232					

Handwritten notes in %D column:
 18
 573
 797
 10383
 20777
 31460
 41232

LB115655

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	18	-		0 SFAM01.1	PPB	8/2/2021 16:09	Std
	0.2 S01	797	-		0.2 SFAM01.1	PPB	8/2/2021 16:12	Std
	2.5 S02	10383	-		2.5 SFAM01.1	PPB	8/2/2021 16:14	Std
	5 S03	20777	-		5 SFAM01.1	PPB	8/2/2021 16:16	Std
	7.5 S04	31460	-		7.5 SFAM01.1	PPB	8/2/2021 16:18	Std
	10 S05	41232	-		10 SFAM01.1	PPB	8/2/2021 16:21	Std
ICV055	ICV055	16515	3.9758 -		SFAM01.1	PPB	8/2/2021 16:23	SMPL
ICB055	ICB055	-71	-0.0248 -		SFAM01.1	PPB	8/2/2021 16:26	SMPL
CCV072	CCV072	20856	5.0229 -		SFAM01.1	PPB	8/2/2021 16:28	SMPL
CCB072	CCB072	-44	-0.0183 -		SFAM01.1	PPB	8/2/2021 16:30	SMPL
PB138120BL	PBW120	-1	-0.008 -		SFAM01.1	PPB	8/2/2021 16:32	SMPL
PB138120TB	LEB120	36	0.001 -		SFAM01.1	PPB	8/2/2021 16:35	SMPL
M3232-05	BG2X1	390	0.0864 -		SFAM01.1	PPB	8/2/2021 16:37	SMPL
M3232-06	BG2X2	28	-0.001 -		SFAM01.1	PPB	8/2/2021 16:39	SMPL
M3232-07	BG2X2D	41	0.0022 -		SFAM01.1	PPB	8/2/2021 16:41	SMPL
M3232-08	BG2X2S	4195	1.0041 -		SFAM01.1	PPB	8/2/2021 16:44	SMPL
CCV073	CCV073	21125	5.0878 -		SFAM01.1	PPB	8/2/2021 16:46	SMPL
CCB073	CCB073	-32	-0.0154 -		SFAM01.1	PPB	8/2/2021 16:48	SMPL