



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115671

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/03/2021	09:05
2	M3242-01		1	8	100.00	>212.0	>212.0	08/03/2021	09:35
3	M3242-01DUP		1	8	100.00	>212.0	>212.0	08/03/2021	09:58
4	M3257-01		1	9	100.00	>212.0	>212.0	08/03/2021	10:20
5	M3258-01		1	9	100.00	>212.0	>212.0	08/03/2021	10:45
6	M3258-03		1	10	100.00	>212.0	>212.0	08/03/2021	11:05
7	M3259-25		1	9	100.00	>212.0	>212.0	08/03/2021	11:30
8	M3259-27		1	10	100.00	>212.0	>212.0	08/03/2021	11:50
9	M3259-29		1	9	100.00	>212.0	>212.0	08/03/2021	12:10
10	M3259-31		1	10	100.00	>212.0	>212.0	08/03/2021	12:32
11	M3259-33		1	9	100.00	>212.0	>212.0	08/03/2021	12:55
12	M3260-01		1	9	100.00	>212.0	>212.0	08/03/2021	13:15
13	M3260-03		1	10	100.00	>212.0	>212.0	08/03/2021	13:40
14	M3261-01		1	10	100.00	>212.0	>212.0	08/03/2021	14:02
15	M3261-03		1	9	100.00	>212.0	>212.0	08/03/2021	14:25
16	M3261-05		1	10	100.00	>212.0	>212.0	08/03/2021	14:45
17	M3266-01		1	10	100.00	>212.0	>212.0	08/03/2021	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB115671

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8321

WorkList ID : 151128

Department : Wet-Chemistry

Date : 08-03-2021 08:35:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2021	Water	M3242-01	Flash Point	Cool 4 deg C	PSEG03	D21	257392	07/30/2021	1010B
08/09/2021	Water	M3257-01	Flash Point	Cool 4 deg C	PSEG03	D22	33251	08/02/2021	1010B
08/09/2021	Water	M3258-01	Flash Point	Cool 4 deg C	PSEG03	D22	33377	08/02/2021	1010B
08/09/2021	Water	M3258-03	Flash Point	Cool 4 deg C	PSEG03	D22	33378	08/02/2021	1010B
08/05/2021	Water	M3259-25	Flash Point	Cool 4 deg C	PSEG03	D22	TOTE-COMP	08/02/2021	1010B
08/05/2021	Water	M3259-27	Flash Point	Cool 4 deg C	PSEG03	D22	X-COMP	08/02/2021	1010B
08/05/2021	Water	M3259-29	Flash Point	Cool 4 deg C	PSEG03	D22	B-COMP	08/02/2021	1010B
08/05/2021	Water	M3259-31	Flash Point	Cool 4 deg C	PSEG03	D22	A-COMP	08/02/2021	1010B
08/05/2021	Water	M3259-33	Flash Point	Cool 4 deg C	PSEG03	D22	O-COMP	08/02/2021	1010B
08/09/2021	Water	M3260-01	Flash Point	Cool 4 deg C	PSEG03	D22	33368	08/02/2021	1010B
08/09/2021	Water	M3260-03	Flash Point	Cool 4 deg C	PSEG03	D22	33389	08/02/2021	1010B
08/05/2021	Water	M3261-01	Flash Point	Cool 4 deg C	PSEG03	D22	31102	08/02/2021	1010B
08/05/2021	Water	M3261-03	Flash Point	Cool 4 deg C	PSEG03	D22	41936	08/02/2021	1010B
08/05/2021	Water	M3261-05	Flash Point	Cool 4 deg C	PSEG03	D22	41937-38	08/02/2021	1010B
08/16/2021	Water	M3266-01	Flash Point	Cool 4 deg C	PSEG03	D22	33388	08/02/2021	1010B

Date/Time 08/03/21 08:40
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 08/03/21 15:15
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: KS (WC)