

LB 115677


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

52488



A= 0.0000e+000

B= 1.9147e-004

C= 1.6756e-002

Rho= 0.9999179

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.032	0.032	79	0.000	79					-
0.2	0.200	0.239	0.039	1163	0.0 %	1163					20
2.5	2.500	2.456	-0.044	12739	0.0 %	12739					-2
5.0	5.000	4.949	-0.051	25762	0.0 %	25762					-1
7.5	7.500	7.457	-0.043	38856	0.0 %	38856					-1
10.0	10.000	10.067	0.067	52488	0.0 %	52488					1

LB115677 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Aug 2021 15:18:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	79	-	0.00007471B	PPB		03 Aug 2021 15:35:45	S	Std
0.2 - 1	50.2	1163	-	0.20007471B	PPB		03 Aug 2021 15:38:02	S	Std
2.5 - 1	52.5	12739	-	2.50007471B	PPB		03 Aug 2021 15:40:19	S	Std
5.0 - 1	55	25762	-	5.00007471B	PPB		03 Aug 2021 15:42:36	S	Std
7.5 - 1	57.5	38856	-	7.50007471B	PPB		03 Aug 2021 15:44:52	S	Std
10.0 - 1	510	52488	-	10.00007471B	PPB		03 Aug 2021 15:47:07	S	Std
ICV35 - 1	ICV35	19932	3.8332	-7471B	PPB		03 Aug 2021 15:49:53	U	SMPL
ICB35 - 1	ICB35	-12	0.0145	-7471B	PPB		03 Aug 2021 15:52:07	U	SMPL
CCV78 - 1	CCV78	25968	4.9889	-7471B	PPB		03 Aug 2021 15:54:23	U	SMPL
CCB78 - 1	CCB78	-8	0.0152	-7471B	PPB		03 Aug 2021 15:56:37	U	SMPL
CRA - 1	CRA	1151	0.2371	-7471B	PPB		03 Aug 2021 15:58:52	U	SMPL
HighStd - 1	HighStd	51887	9.9517	-7471B	PPB		03 Aug 2021 16:01:07	U	SMPL
ChkStd - 1	ChkStd	37368	7.1717	-7471B	PPB		03 Aug 2021 16:03:23	U	SMPL
PB138167BL - 1	PBS	-114	-0.0051	-7471B	PPB		03 Aug 2021 16:05:41	U	SMPL
PB138167BS - 1	LCSS	21285	4.0922	-7471B	PPB		03 Aug 2021 16:07:58	U	SMPL
M3241-01 - 1	20210583-COM-6	41355	7.9351	-7471B	PPB		03 Aug 2021 16:10:14	U	SMPL
M3241-02 - 1	20210592-AMENDED-COM-6	43271	8.3019	-7471B	PPB		03 Aug 2021 16:12:31	U	SMPL
M3253-01 - 1	MH-1-COMP	830	0.1757	-7471B	PPB		03 Aug 2021 16:14:49	U	SMPL
M3253-03 - 1	MH-2-COMP	5040	0.9818	-7471B	PPB		03 Aug 2021 16:17:06	U	SMPL
M3255-01 - 1	FRAC-TANK-COMP	10259	1.9811	-7471B	PPB		03 Aug 2021 16:19:21	U	SMPL
CCV79 - 1	CCV79	25955	4.9864	-7471B	PPB		03 Aug 2021 16:21:36	U	SMPL
CCB79 - 1	CCB79	10	0.0187	-7471B	PPB		03 Aug 2021 16:23:51	U	SMPL
M3256-01 - 1	OK-03-080221	946	0.1979	-7471B	PPB		03 Aug 2021 16:26:06	U	SMPL
M3256-01DUP - 1	OK-03-080221DUP	886	0.1864	-7471B	PPB		03 Aug 2021 16:28:22	U	SMPL
M3256-01MS - 1	OK-03-080221MS	22446	4.3145	-7471B	PPB		03 Aug 2021 16:30:37	U	SMPL
M3256-01MSD - 1	OK-03-080221MSD	22232	4.2736	-7471B	PPB		03 Aug 2021 16:32:52	U	SMPL
M3256-01LX5 - 1		196	0.0543	-7471B	PPB		03 Aug 2021 16:35:08	U	SMPL
M3256-01A - 1		22167	4.2611	-7471B	PPB		03 Aug 2021 16:37:24	U	SMPL
CCV80 - 1	CCV80	25912	4.9782	-7471B	PPB		03 Aug 2021 16:39:41	U	SMPL
CCB80 - 1	CCB80	6	0.0179	-7471B	PPB		03 Aug 2021 16:41:57	U	SMPL