



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115678

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/03/2021	15:30
2	CAL2	1	Water	NA	NA	21.2	7.00	08/03/2021	15:31
3	CAL3	1	Water	NA	NA	21.2	10.03	08/03/2021	15:33
4	ICV	1	Water	NA	NA	21.3	7.00	08/03/2021	15:35
5	CCV1	1	Water	NA	NA	21.2	2.01	08/03/2021	15:36
6	M3276-02	1	Solid	20.03	20	22.7	8.25	08/03/2021	15:39
7	M3276-02DUP	1	Solid	20.04	20	22.8	8.26	08/03/2021	15:40
8	M3276-04	1	Solid	20.05	20	22.6	7.94	08/03/2021	15:42
9	CCV2	1	Water	NA	NA	21.3	12.02	08/03/2021	15:45

18115678

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M3276 WorkList ID : 151134 Department : Wet-Chemistry Date : 08-03-2021 15:11:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/02/2021	Solid	M3276-02	Corrosivity	Cool 4 deg C	NOBI03	L11	PRE-MAEC-01	08/02/2021	9045D
08/02/2021	Solid	M3276-04	Corrosivity	Cool 4 deg C	NOBI03	L11	PRE-MAEC-02	08/02/2021	9045D

Date/Time 08/03/21 15:15
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: CP SM

Date/Time 08/03/21 16:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WC)