


LB115683

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5616e-004

C= 5.6182e-003

Rho= 0.9998866

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.008	0.008	8	0.000	8					-
0.2	0.200	0.214	0.014	812	0.0 %	812					7
2.5	2.500	2.462	-0.038	9588	0.0 %	9588					-2
5.0	5.000	4.964	-0.036	19356	0.0 %	19356					-1
7.5	7.500	7.611	0.111	29688	0.0 %	29688					1
10.0	10.000	9.943	-0.057	38792	0.0 %	38792					-1

LB115683 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Aug 2021 10:33:41				
Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type
0.0 - 1	50	8	-	0.0007470A	PPB		04 Aug 2021 10:35:23	S Std
0.2 - 1	50.2	812	-	0.20007470A	PPB		04 Aug 2021 10:37:39	S Std
2.5 - 1	52.5	9588	-	2.50007470A	PPB		04 Aug 2021 10:39:56	S Std
5.0 - 1	55	19356	-	5.00007470A	PPB		04 Aug 2021 10:42:13	S Std
7.5 - 1	57.5	29688	-	7.50007470A	PPB		04 Aug 2021 10:44:29	S Std
10.0 - 1	510	38792	-	10.00007470A	PPB		04 Aug 2021 10:49:18	S Std
ICV36 - 1	ICV36	15232	3.9075	-7470A	PPB		04 Aug 2021 10:51:58	U SMPL
ICB36 - 1	ICB36	-43	-0.0054	-7470A	PPB		04 Aug 2021 10:54:13	U SMPL
CCV81 - 1	CCV81	19594	5.0248	-7470A	PPB		04 Aug 2021 10:56:27	U SMPL
CCB81 - 1	CCB81	-57	-0.0090	-7470A	PPB		04 Aug 2021 10:58:42	U SMPL
CRA - 1	CRA	808	0.2126	-7470A	PPB		04 Aug 2021 11:00:58	U SMPL
HighStd - 1	HighStd	39120	10.0266	-7470A	PPB		04 Aug 2021 11:03:13	U SMPL
ChkStd - 1	ChkStd	26598	6.8190	-7470A	PPB		04 Aug 2021 11:05:27	U SMPL
PB138163BL - 1	PBW	-152	-0.0333	-7470A	PPB		04 Aug 2021 11:07:44	U SMPL
PB138163BS - 1	LCSW	13533	3.4722	-7470A	PPB		04 Aug 2021 11:10:01	U SMPL
M3051-14 - 1	41213	-63	-0.0105	-7470A	PPB		04 Aug 2021 11:12:17	U SMPL
M3255-02 - 1	FRAC- TANK-COMP	-7	0.0038	-7470A	PPB		04 Aug 2021 11:14:33	U SMPL
M3255-02DUP - 1	FRAC- TANK-COMPDUP	24	0.0118	-7470A	PPB		04 Aug 2021 11:16:50	U SMPL
M3255-02MS - 1	FRAC- TANK-COMPMS	13179	3.3816	-7470A	PPB		04 Aug 2021 11:19:06	U SMPL
M3255-02MSD - 1	FRAC- TANK-COMPMSD	13408	3.4402	-7470A	PPB		04 Aug 2021 11:21:22	U SMPL
CCV82 - 1	CCV82	19463	4.9913	-7470A	PPB		04 Aug 2021 11:23:37	U SMPL
CCB82 - 1	CCB82	-49	-0.0069	-7470A	PPB		04 Aug 2021 11:25:53	U SMPL
PB138164BL - 1	PBW	-11	0.0028	-7470A	PPB		04 Aug 2021 11:28:08	U SMPL
PB138164BS - 1	LCSW	12564	3.2240	-7470A	PPB		04 Aug 2021 11:30:23	U SMPL
M3241-03 - 1	SPLP-AMENDED-COM-6	155	0.0453	-7470A	PPB		04 Aug 2021 11:32:38	U SMPL
M3241-03DUP - 1	SPLP-AMENDED-COM-6DUP	167	0.0484	-7470A	PPB		04 Aug 2021 11:34:53	U SMPL
M3241-03MS - 1	SPLP-AMENDED-COM-6MS	15341	3.9354	-7470A	PPB		04 Aug 2021 11:40:05	U SMPL
M3241-03MSD - 1	SPLP-AMENDED-COM-6MSD	15214	3.9029	-7470A	PPB		04 Aug 2021 11:42:21	U SMPL
PB138165BL - 1	PBW	-89	-0.0172	-7470A	PPB		04 Aug 2021 11:44:37	U SMPL
PB138165BS - 1	LCSW	12579	3.2279	-7470A	PPB		04 Aug 2021 11:46:54	U SMPL
M3261-01 - 1	31102	81086	20.7767	-7470A	PPB		04 Aug 2021 11:49:11	U SMPL
M3261-01DUP - 1	31102DUP	91259	23.3826	-7470A	PPB		04 Aug 2021 11:51:27	U SMPL
CCV83 - 1	CCV83	19419	4.9800	-7470A	PPB		04 Aug 2021 11:53:43	U SMPL
CCB83 - 1	CCB83	-44	-0.0057	-7470A	PPB		04 Aug 2021 11:55:59	U SMPL
M3261-01MS - 1	31102MS	98638	25.2728	-7470A	PPB		04 Aug 2021 11:58:15	U SMPL
M3261-01MSD - 1	31102MSD	97146	24.8906	-7470A	PPB		04 Aug 2021 12:00:30	U SMPL
M3261-03 - 1	41936	1102	0.2879	-7470A	PPB		04 Aug 2021 12:02:47	U SMPL
M3261-05 - 1	41937-38	12	0.0087	-7470A	PPB		04 Aug 2021 12:05:04	U SMPL
PB135163TB - 1	PB138130TB	1	0.0059	-7470A	PPB		04 Aug 2021 12:07:20	U SMPL
PB135164TB - 1	PB138158TB	-30	-0.0021	-7470A	PPB		04 Aug 2021 12:09:36	U SMPL
CCV84 - 1	CCV84	20411	5.2341	-7470A	PPB		04 Aug 2021 12:11:53	U SMPL
CCB84 - 1	CCB84	-17	0.0013	-7470A	PPB		04 Aug 2021 12:14:10	U SMPL
M3255-02LX5 - 1		-48	-0.0067	-7470A	PPB		04 Aug 2021 12:16:28	U SMPL
M3255-02A - 1		16934	4.3435	-7470A	PPB		04 Aug 2021 12:18:43	U SMPL
M3241-03LX5 - 1		-21	0.0002	-7470A	PPB		04 Aug 2021 12:20:59	U SMPL
M3241-03A - 1		16572	4.2507	-7470A	PPB		04 Aug 2021 12:23:14	U SMPL
M3261-01LX5 - 1		16975	4.3540	-7470A	PPB		04 Aug 2021 12:25:30	U SMPL
M3261-01A - 1		102216	26.1894	-7470A	PPB		04 Aug 2021 12:27:45	U SMPL
CCV85 - 1	CCV85	19190	4.9213	-7470A	PPB		04 Aug 2021 12:30:01	U SMPL
CCB85 - 1	CCB85	-17	0.0013	-7470A	PPB		04 Aug 2021 12:32:17	U SMPL
M3261-01DLX5 - 1		16225	4.1618	-7470A	PPB		04 Aug 2021 12:34:33	U SMPL
M3261-01DUPDLX5 - 1		18805	4.8227	-7470A	PPB		04 Aug 2021 12:36:49	U SMPL
M3261-01MSDLX5 - 1		20334	5.2144	-7470A	PPB		04 Aug 2021 12:39:06	U SMPL
M3261-01MSDDLX5 - 1		19758	5.0668	-7470A	PPB		04 Aug 2021 12:41:23	U SMPL
M3261-01LDLX25 - 1		3064	0.7905	-7470A	PPB		04 Aug 2021 12:43:39	U SMPL
M3261-01ADLX5 - 1		21269	5.4539	-7470A	PPB		04 Aug 2021 12:45:55	U SMPL
CCV86 - 1	CCV86	20201	5.1803	-7470A	PPB		04 Aug 2021 12:48:10	U SMPL
CCB86 - 1	CCB86	-23	-0.0003	-7470A	PPB		04 Aug 2021 12:50:26	U SMPL