



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115691

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/04/2021	12:15
2	CAL2	1	Water	NA	NA	20.3	6.99	08/04/2021	12:16
3	CAL3	1	Water	NA	NA	20.3	10.03	08/04/2021	12:17
4	ICV	1	Water	NA	NA	20.4	7.01	08/04/2021	12:19
5	CCV1	1	Water	NA	NA	20.4	2.01	08/04/2021	12:20
6	M3283-02	1	Water	NA	NA	21.1	7.84	08/04/2021	12:22
7	M3283-02DUP	1	Water	NA	NA	21.2	7.85	08/04/2021	12:23
8	M3283-03	1	Water	NA	NA	21.3	7.29	08/04/2021	12:25
9	CCV2	1	Water	NA	NA	20.4	12.02	08/04/2021	12:26

WORKLIST(Hardcopy Internal Chain)

13115691

WorkList Name : ph m3283 WorkList ID : 151150 Department : Wet-Chemistry Date : 08-04-2021 12:03:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/03/2021	Water	M3283-02	pH	Cool 4 deg C	TETR06	L11	TT-031-IDWGW-2021080	08/03/2021	9040C
08/03/2021	Water	M3283-03	pH	Cool 4 deg C	TETR06	L11	TT-032-IDWGW-2021080	08/03/2021	9040C

Date/Time 08/04/21 12:10
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: Z-C (San)

Date/Time 08/04/21 13:00
 Raw Sample Received by: Z-C (San)
 Raw Sample Relinquished by: JP (WC)