



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115697

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/04/2021	14:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/04/2021	14:16
3	CAL3	1	Water	NA	NA	20.6	10.03	08/04/2021	14:18
4	ICV	1	Water	NA	NA	20.6	7.01	08/04/2021	14:20
5	CCV1	1	Water	NA	NA	20.6	2.01	08/04/2021	14:22
6	M3281-02	1	Solid	20.02	20	22.5	10.74	08/04/2021	14:25
7	M3281-02DUP	1	Solid	20.03	20	22.6	10.75	08/04/2021	14:26
8	M3291-02	1	Solid	20.01	20	23.3	8.30	08/04/2021	14:29
9	M3291-05	1	Solid	20.04	20	23.3	7.46	08/04/2021	14:30
10	CCV2	1	Water	NA	NA	20.7	12.02	08/04/2021	14:32

WORKLIST(Hardcopy Internal Chain)

WB 115697

WorkList Name : CORROSIONITY M3291 WorkList ID : 151161 Department : Wet-Chemistry Date : 08-04-2021 14:01:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/03/2021	Solid	M3281-02	Corrosivity	Cool 4 deg C	loui01	D31	TWP03	08/03/2021	9045D
08/04/2021	Solid	M3291-02	Corrosivity	Cool 4 deg C	PSEG03	D31	WC-1	08/04/2021	9045D
08/04/2021	Solid	M3291-05	Corrosivity	Cool 4 deg C	PSEG03	D31	WC-2	08/04/2021	9045D

Date/Time 08/04/21 14:10
 Raw Sample Received by: JP (W9)
 Raw Sample Relinquished by: JP (SW)

Date/Time 08/04/21 16:30
 Raw Sample Received by: JP (SW)
 Raw Sample Relinquished by: JP (WC)