


LB115700

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5757e-004

C= 2.2273e-002

Rho= 0.9997898

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.032	0.032	36	0.000	36					-
0.2	0.200	0.229	0.029	803	0.0 %	803					15
2.5	2.500	2.533	0.033	9749	0.0 %	9749					1
5.0	5.000	4.868	-0.132	18811	0.0 %	18811					-3
7.5	7.500	7.438	-0.062	28792	0.0 %	28792					-1
10.0	10.000	10.100	0.100	39126	0.0 %	39126					1

LB115700 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Aug 2021 08:09:52

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SD	36	-	0.00007470A	PPB	05 Aug 2021 08:21:29	S	Std
0.2 - 1	50.2	803	-	0.20007470A	PPB	05 Aug 2021 08:23:45	S	Std
2.5 - 1	52.5	9749	-	2.50007470A	PPB	05 Aug 2021 08:26:02	S	Std
5.0 - 1	55	18811	-	5.00007470A	PPB	05 Aug 2021 08:28:19	S	Std
7.5 - 1	57.5	28792	-	7.50007470A	PPB	05 Aug 2021 08:30:35	S	Std
10.0 - 1	59	39126	-	10.00007470A	PPB	05 Aug 2021 08:38:00	S	Std
ICV38 - 1	ICV38	15338	3.9729	-7470A	PPB	05 Aug 2021 08:40:45	U	SMPL
ICB38 - 1	ICB38	-60	0.0068	-7470A	PPB	05 Aug 2021 08:43:00	U	SMPL
CCV90 - 1	CCV90	18253	4.7238	-7470A	PPB	05 Aug 2021 08:45:16	U	SMPL
CCB90 - 1	CCB90	-59	0.0071	-7470A	PPB	05 Aug 2021 08:47:32	U	SMPL
CRA - 1	CRA	724	0.2088	-7470A	PPB	05 Aug 2021 08:49:47	U	SMPL
HighStd - 1	HighStd	38597	9.9639	-7470A	PPB	05 Aug 2021 08:52:02	U	SMPL
ChkStd - 1	ChkStd	25913	6.6968	-7470A	PPB	05 Aug 2021 08:54:18	U	SMPL
PB138192BL - 1	PBW	-104	-0.0045	-7470A	PPB	05 Aug 2021 08:56:34	U	SMPL
PB138192BS - 1	LCSW	14078	3.6484	-7470A	PPB	05 Aug 2021 08:58:50	U	SMPL
M3271-03 - 1	TP-1	-48	0.0099	-7470A	PPB	05 Aug 2021 09:01:09	U	SMPL
M3271-06 - 1	TP-2	7	0.0241	-7470A	PPB	05 Aug 2021 09:03:29	U	SMPL
M3271-09 - 1	TP-3	29	0.0297	-7470A	PPB	05 Aug 2021 09:05:46	U	SMPL
M3271-12 - 1	TP-4	20	0.0274	-7470A	PPB	05 Aug 2021 09:08:03	U	SMPL
M3271-15 - 1	TP-5	30	0.0300	-7470A	PPB	05 Aug 2021 09:10:18	U	SMPL
CCV91 - 1	CCV91	18856	4.8791	-7470A	PPB	05 Aug 2021 09:12:33	U	SMPL
CCB91 - 1	CCB91	-57	0.0076	-7470A	PPB	05 Aug 2021 09:14:48	U	SMPL
M3276-02 - 1	PRE-MAEC-01	19	0.0272	-7470A	PPB	05 Aug 2021 09:17:03	U	SMPL
M3276-04 - 1	PRE-MAEC-02	10	0.0248	-7470A	PPB	05 Aug 2021 09:19:18	U	SMPL
M3276-04DUP - 1	PRE-MAEC-02DUP	16	0.0264	-7470A	PPB	05 Aug 2021 09:21:33	U	SMPL
M3276-04MS - 1	PRE-MAEC-02MS	14246	3.6917	-7470A	PPB	05 Aug 2021 09:23:49	U	SMPL
M3276-04MSD - 1	PRE-MAEC-02MSD	14455	3.7455	-7470A	PPB	05 Aug 2021 09:26:04	U	SMPL
PB138201BL - 1	PBW	-26	0.0156	-7470A	PPB	05 Aug 2021 09:28:20	U	SMPL
PB138201BS - 1	LCSW	14104	3.6551	-7470A	PPB	05 Aug 2021 09:30:37	U	SMPL
M3283-02 - 1	TT-031-IDWGW-20210803	671	0.1951	-7470A	PPB	05 Aug 2021 09:32:53	U	SMPL
M3283-03 - 1	TT-032-IDWGW-20210803	47	0.0344	-7470A	PPB	05 Aug 2021 09:35:10	U	SMPL
M3283-03DUP - 1	TT-032-IDWGW-20210803DUP	64	0.0388	-7470A	PPB	05 Aug 2021 09:37:47	U	SMPL
CCV92 - 1	CCV92	19018	4.9208	-7470A	PPB	05 Aug 2021 09:40:02	U	SMPL
CCB92 - 1	CCB92	-42	0.0115	-7470A	PPB	05 Aug 2021 09:42:18	U	SMPL
M3283-03MS - 1	TT-032-IDWGW-20210803MS	18125	4.6908	-7470A	PPB	05 Aug 2021 09:44:34	U	SMPL
M3283-03MSD - 1	TT-032-IDWGW-20210803MSD	16345	4.2323	-7470A	PPB	05 Aug 2021 09:46:49	U	SMPL
PB138192TB - 1	PB138171TB	-22	0.0166	-7470A	PPB	05 Aug 2021 09:49:04	U	SMPL
M3276-04LX5 - 1		18	0.0269	-7470A	PPB	05 Aug 2021 09:51:20	U	SMPL
M3276-04A - 1		15512	4.0178	-7470A	PPB	05 Aug 2021 09:53:36	U	SMPL
M3283-03LX5 - 1		-28	0.0151	-7470A	PPB	05 Aug 2021 09:55:52	U	SMPL
M3283-03A - 1		15445	4.0005	-7470A	PPB	05 Aug 2021 09:58:08	U	SMPL
CCV93 - 1	CCV93	19073	4.9350	-7470A	PPB	05 Aug 2021 10:00:24	U	SMPL
CCB93 - 1	CCB93	-30	0.0145	-7470A	PPB	05 Aug 2021 10:02:42	U	SMPL