



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115703

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2810         |
| buffer solution pH 7 yellow    | W2813         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2809         |
| buffer solution pH 7 yellow    | W2719         |
| Buffer Solution, PH2 (500ml)   | W2806         |
| Buffer Solution, PH12 (500ml)  | W2807         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 08/05/2021 | 11:15     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 08/05/2021 | 11:16     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.02       | 08/05/2021 | 11:18     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.3             | 7.01        | 08/05/2021 | 11:20     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 08/05/2021 | 11:22     |
| 6   | M3293-05    | 1  | Solid  | 20.01       | 20          | 21.3             | 5.06        | 08/05/2021 | 11:25     |
| 7   | M3293-05DUP | 1  | Solid  | 20.02       | 20          | 21.4             | 5.07        | 08/05/2021 | 11:26     |
| 8   | M3298-01    | 1  | Solid  | 20.03       | 20          | 20.7             | 5.05        | 08/05/2021 | 11:30     |
| 9   | M3298-02    | 1  | Solid  | 20.04       | 20          | 20.9             | 5.12        | 08/05/2021 | 11:32     |
| 10  | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 08/05/2021 | 11:34     |