



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115704

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/05/2021	11:15
2	CAL2	1	Water	NA	NA	20.2	7.00	08/05/2021	11:16
3	CAL3	1	Water	NA	NA	20.2	10.02	08/05/2021	11:18
4	ICV	1	Water	NA	NA	20.3	7.01	08/05/2021	11:20
5	CCV1	1	Water	NA	NA	20.2	2.01	08/05/2021	11:22
6	M3289-15	1	Water	NA	NA	20.2	7.12	08/05/2021	11:36
7	M3289-16	1	Water	NA	NA	20.1	7.11	08/05/2021	11:37
8	M3289-17	1	Water	NA	NA	20.1	7.07	08/05/2021	11:39
9	M3289-18	1	Water	NA	NA	20.4	7.01	08/05/2021	11:40
10	M3293-09	1	Water	NA	NA	20.1	6.31	08/05/2021	11:42
11	M3293-11	1	Water	NA	NA	20.2	6.46	08/05/2021	11:45
12	M3293-13	1	Water	NA	NA	20.3	9.72	08/05/2021	11:47
13	M3297-03	1	Water	NA	NA	20.3	7.07	08/05/2021	11:50
14	M3297-05	1	Water	NA	NA	20.5	8.34	08/05/2021	11:52
15	M3298-03	1	Water	NA	NA	20.1	7.99	08/05/2021	12:00
16	CCV2	1	Water	NA	NA	20.3	12.02	08/05/2021	12:01
17	M3298-04	1	Water	NA	NA	20.2	6.75	08/05/2021	12:05
18	M3298-05	1	Water	NA	NA	20.4	6.54	08/05/2021	12:09
19	M3298-06	1	Water	NA	NA	20.6	6.83	08/05/2021	12:10
20	M3298-07	1	Water	NA	NA	20.6	6.74	08/05/2021	12:15

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M3301-01	1	Water	NA	NA	22.6	7.21	08/05/2021	12:20
22	M3301-01DUP	1	Water	NA	NA	22.7	7.22	08/05/2021	12:21
23	CCV3	1	Water	NA	NA	20.3	2.01	08/05/2021	12:24

WORKLIST(Hardcopy Internal Chain)

18115704

WorkList Name : PH WATER M3297

WorkList ID : 151193

Department : Wet-Chemistry

Date : 08-05-2021 11:12:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/04/2021	Water	M3289-15	pH	Cool 4 deg C	PSEG03	D31	TOTE-A	08/04/2021	9040C
08/04/2021	Water	M3289-16	pH	Cool 4 deg C	PSEG03	D31	TOTE-B	08/04/2021	9040C
08/04/2021	Water	M3289-17	pH	Cool 4 deg C	PSEG03	D31	TOTE-C	08/04/2021	9040C
08/04/2021	Water	M3289-18	pH	Cool 4 deg C	PSEG03	D31	TOTE-COMP	08/04/2021	9040C
08/04/2021	Water	M3293-09	pH	Cool 4 deg C	PSEG03	D31	TOTES-1-2	08/04/2021	9040C
08/04/2021	Water	M3293-11	pH	Cool 4 deg C	PSEG03	D31	2055-57	08/04/2021	9040C
08/04/2021	Water	M3293-13	pH	Cool 4 deg C	PSEG03	D31	2024-29	08/04/2021	9040C
08/04/2021	Water	M3297-03	pH	Cool 4 deg C	PSEG03	D31	123-24	08/04/2021	9040C
08/04/2021	Water	M3297-05	pH	Cool 4 deg C	PSEG03	D31	121	08/04/2021	9040C
08/04/2021	Water	M3298-03	pH	Cool 4 deg C	PSEG03	D31	1917	08/04/2021	9040C
08/04/2021	Water	M3298-04	pH	Cool 4 deg C	PSEG03	D31	1924	08/04/2021	9040C
08/04/2021	Water	M3298-05	pH	Cool 4 deg C	PSEG03	D31	1928-1930	08/04/2021	9040C
08/04/2021	Water	M3298-06	pH	Cool 4 deg C	PSEG03	D31	1940	08/04/2021	9040C
08/04/2021	Water	M3298-07	pH	Cool 4 deg C	PSEG03	D31	1964-1968	08/04/2021	9040C
08/05/2021	Water	M3301-01	pH	Cool 4 deg C	PSEG03	D32	MH-P17	08/05/2021	9040C

11:15

Date/Time 08/05/21
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 08/05/21
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)