



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115719

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/05/2021	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	08/05/2021	16:01
3	CAL3	1	Water	NA	NA	20.3	10.03	08/05/2021	16:02
4	ICV	1	Water	NA	NA	20.4	7.00	08/05/2021	16:05
5	CCV1	1	Water	NA	NA	20.4	2.01	08/05/2021	16:07
6	M3308-05	1	Solid	20.02	20	22.9	8.62	08/05/2021	16:10
7	M3308-05DUP	1	Solid	20.03	20	22.8	8.63	08/05/2021	16:11
8	CCV2	1	Water	NA	NA	20.5	12.02	08/05/2021	16:12

WORKLIST(Hardcopy Internal Chain)

VB 115719

WorkList Name : CORROSIONITY M3308

WorkList ID : 151205

Department : Wet-Chemistry

Date : 08-05-2021 15:48:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2021	Solid	M3308-05	Corrosivity	Cool 4 deg C	PSEG03	D32	72-11930	08/05/2021	9045D

Date/Time 08/05/21 15:50
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 08/05/21 17:20
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)