



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115720

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/05/2021	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	08/05/2021	16:01
3	CAL3	1	Water	NA	NA	20.3	10.03	08/05/2021	16:02
4	ICV	1	Water	NA	NA	20.4	7.00	08/05/2021	16:05
5	CCV1	1	Water	NA	NA	20.4	2.01	08/05/2021	16:07
6	M3311-01	1	Water	NA	NA	23.6	4.70	08/05/2021	16:15
7	M3311-01DUP	1	Water	NA	NA	23.7	4.71	08/05/2021	16:16
8	CCV2	1	Water	NA	NA	20.6	12.03	08/05/2021	16:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3311w WorkList ID : 151204 Department : Wet-Chemistry Date : 08-05-2021 15:48:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/03/2021	Water	M3311-01	pH	Cool 4 deg C	STER04	L11	POOL-WATER	08/03/2021	9040C

Date/Time 08/05/21 15:50
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: CR SM

Date/Time 08/05/21 17:20
 Raw Sample Received by: CR SM
 Raw Sample Relinquished by: JR (WC)