


LB115728

7470A

INSTRUMENT ID: 2V1



Linear

A= 0.0000e+000

B= 2.3575e-004

C= 4.2407e-002

Rho= 0.9987778

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.042	0.042	0	0.000	0					
0.2	0.200	0.207	0.007	698	0.0 %	698					3
2.5	2.500	2.563	0.063	10691	0.0 %	10691					3
5.0	5.000	4.697	-0.303	19744	0.0 %	19744					-6
7.5	7.500	7.800	0.300	32906	0.0 %	32906					4
10.0	10.000	9.891	-0.109	41773	0.0 %	41773					-1

LB115728 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 06 Aug 2021 10:51:41					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7470A	PPB	06 Aug 2021 10:55:38	S	Std
0.2 - 1	50.2	698	-	0.2000	7470A	PPB	06 Aug 2021 10:57:55	S	Std
2.5 - 1	52.5	10691	-	2.5000	7470A	PPB	06 Aug 2021 11:00:12	S	Std
5.0 - 1	55	19744	-	5.0000	7470A	PPB	06 Aug 2021 11:02:28	S	Std
7.5 - 1	57.5	32906	-	7.5000	7470A	PPB	06 Aug 2021 11:04:44	S	Std
10.0 - 1	511	41773	-	10.0000	7470A	PPB	06 Aug 2021 11:15:17	S	Std
ICV42 - 1	ICV42	17105	4.0750	-	7470A	PPB	06 Aug 2021 11:20:36	U	SMPL
ICB42 - 1	ICB42	-93	0.0205	-	7470A	PPB	06 Aug 2021 11:22:50	U	SMPL
CCV04 - 1	CCV04	18988	4.5189	-	7470A	PPB	06 Aug 2021 11:25:05	U	SMPL
CCB04 - 1	CCB04	-79	0.0238	-	7470A	PPB	06 Aug 2021 11:27:20	U	SMPL
CRA - 1	CRA	676	0.2018	-	7470A	PPB	06 Aug 2021 11:29:36	U	SMPL
HighStd - 1	HighStd	41551	9.8382	-	7470A	PPB	06 Aug 2021 11:31:50	U	SMPL
ChkStd - 1	ChkStd	30353	7.1982	-	7470A	PPB	06 Aug 2021 11:37:09	U	SMPL
PB138226BL - 1	PBW	-163	0.0040	-	7470A	PPB	06 Aug 2021 11:39:24	U	SMPL
PB138226BS - 1	LCSW	15613	3.7232	-	7470A	PPB	06 Aug 2021 11:44:19	U	SMPL
M3281-02 - 1	TWP03	-109	0.0167	-	7470A	PPB	06 Aug 2021 11:46:37	U	SMPL
M3291-02 - 1	WC-1	-9	0.0403	-	7470A	PPB	06 Aug 2021 11:48:52	U	SMPL
M3291-05 - 1	WC-2	121	0.0709	-	7470A	PPB	06 Aug 2021 11:51:09	U	SMPL
M3291-05DUP - 1	WC-2DUP	26	0.0485	-	7470A	PPB	06 Aug 2021 11:58:13	U	SMPL
M3291-05MS - 1	WC-2MS	15979	3.8095	-	7470A	PPB	06 Aug 2021 12:00:27	U	SMPL
CCV05 - 1	CCV05	21074	5.0107	-	7470A	PPB	06 Aug 2021 12:02:42	U	SMPL
CCB05 - 1	CCB05	-88	0.0217	-	7470A	PPB	06 Aug 2021 12:04:57	U	SMPL
M3291-05MSD - 1	WC-2MSD	16102	3.8385	-	7470A	PPB	06 Aug 2021 12:09:50	U	SMPL
M3298-01 - 1	1949	261	0.1039	-	7470A	PPB	06 Aug 2021 12:12:05	U	SMPL
M3298-02 - 1	1969	11	0.0450	-	7470A	PPB	06 Aug 2021 12:14:20	U	SMPL
PB138227BL - 1	PBW	39	0.0516	-	7470A	PPB	06 Aug 2021 12:16:36	U	SMPL
PB138227BS - 1	LCSW	15195	3.6247	-	7470A	PPB	06 Aug 2021 12:21:23	U	SMPL
M3298-03 - 1	1917	-47	0.0313	-	7470A	PPB	06 Aug 2021 12:23:39	U	SMPL
M3298-04 - 1	1924	37	0.0511	-	7470A	PPB	06 Aug 2021 12:25:57	U	SMPL
M3298-05 - 1	1928-1930	24	0.0481	-	7470A	PPB	06 Aug 2021 12:28:14	U	SMPL
M3298-06 - 1	1940	161	0.0804	-	7470A	PPB	06 Aug 2021 12:30:31	U	SMPL
M3298-07 - 1	1964-1968	36	0.0509	-	7470A	PPB	06 Aug 2021 12:32:46	U	SMPL
CCV06 - 1	CCV06	21178	5.0352	-	7470A	PPB	06 Aug 2021 12:35:02	U	SMPL
CCB06 - 1	CCB06	-74	0.0250	-	7470A	PPB	06 Aug 2021 12:37:17	U	SMPL
M3298-07DUP - 1	1964-1968DUP	18	0.0467	-	7470A	PPB	06 Aug 2021 12:39:33	U	SMPL
M3298-07MS - 1	1964-1968MS	15234	3.6339	-	7470A	PPB	06 Aug 2021 12:52:10	U	SMPL
M3298-07MSD - 1	1964-1968MSD	15958	3.8046	-	7470A	PPB	06 Aug 2021 12:54:27	U	SMPL
M3298-08 - 1	ANTIFREEZE	-45	0.0318	-	7470A	PPB	06 Aug 2021 12:56:43	U	SMPL
PB138228BL - 1	PBW	-10	0.0400	-	7470A	PPB	06 Aug 2021 12:58:58	U	SMPL
PB138228BS - 1	LCSW	16746	3.9903	-	7470A	PPB	06 Aug 2021 13:01:14	U	SMPL
M3289-18 - 1	TOTE-COMP	415	0.1402	-	7470A	PPB	06 Aug 2021 13:03:30	U	SMPL
M3293-09 - 1	TOTES-1-2	-18	0.0382	-	7470A	PPB	06 Aug 2021 13:05:47	U	SMPL
M3293-11 - 1	2055-57	215	0.0931	-	7470A	PPB	06 Aug 2021 13:08:04	U	SMPL
M3293-13 - 1	2024-29	1472	0.3894	-	7470A	PPB	06 Aug 2021 13:10:19	U	SMPL
CCV07 - 1	CCV07	21015	4.9968	-	7470A	PPB	06 Aug 2021 13:12:35	U	SMPL
CCB07 - 1	CCB07	-65	0.0271	-	7470A	PPB	06 Aug 2021 13:14:50	U	SMPL
M3297-02 - 1	125-27	331	0.1204	-	7470A	PPB	06 Aug 2021 13:17:06	U	SMPL
M3297-03 - 1	123-24	-14	0.0391	-	7470A	PPB	06 Aug 2021 13:19:21	U	SMPL
M3297-05 - 1	121	166	0.0815	-	7470A	PPB	06 Aug 2021 13:21:37	U	SMPL
M3301-01 - 1	MH-P17	231	0.0969	-	7470A	PPB	06 Aug 2021 13:23:52	U	SMPL
M3301-01DUP - 1	MH-P17DUP	247	0.1006	-	7470A	PPB	06 Aug 2021 13:26:09	U	SMPL
M3301-01MS - 1	MH-P17MS	15593	3.7185	-	7470A	PPB	06 Aug 2021 13:28:24	U	SMPL
M3301-01MSD - 1	MH-P17MSD	15143	3.6124	-	7470A	PPB	06 Aug 2021 13:30:41	U	SMPL
PB138226TB - 1	PB138197TB	-65	0.0271	-	7470A	PPB	06 Aug 2021 13:32:58	U	SMPL
PB138227TB - 1	PB138224TB	18	0.0467	-	7470A	PPB	06 Aug 2021 13:35:15	U	SMPL
CCV08 - 1	CCV08	20514	4.8787	-	7470A	PPB	06 Aug 2021 13:37:30	U	SMPL
CCB08 - 1	CCB08	-62	0.0278	-	7470A	PPB	06 Aug 2021 13:39:46	U	SMPL
M3291-05LX5 - 1		0	0.0424	-	7470A	PPB	06 Aug 2021 13:42:02	U	SMPL
M3291-05A - 1		15611	3.7228	-	7470A	PPB	06 Aug 2021 13:44:17	U	SMPL
M3298-07LX5 - 1		-91	0.0210	-	7470A	PPB	06 Aug 2021 13:46:33	U	SMPL
M3298-07A - 1		16671	3.9727	-	7470A	PPB	06 Aug 2021 13:48:48	U	SMPL
M3301-01LX5 - 1		-43	0.0323	-	7470A	PPB	06 Aug 2021 13:51:03	U	SMPL
M3301-01A - 1		16657	3.9694	-	7470A	PPB	06 Aug 2021 13:53:19	U	SMPL
CCV09 - 1	CCV09	20942	4.9796	-	7470A	PPB	06 Aug 2021 13:55:35	U	SMPL
CCB09 - 1	CCB09	-72	0.0254	-	7470A	PPB	06 Aug 2021 13:57:52	U	SMPL