



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115731

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	08/06/2021	12:05
2	M3289-18		1	9	100.00	>212.0	>212.0	08/06/2021	12:35
3	M3289-18DUP		1	9	100.00	>212.0	>212.0	08/06/2021	12:58
4	M3293-09		1	10	100.00	>212.0	>212.0	08/06/2021	13:20
5	M3293-11		1	10	100.00	>212.0	>212.0	08/06/2021	13:45
6	M3293-13		1	9	100.00	>212.0	>212.0	08/06/2021	14:10
7	M3301-01		1	10	100.00	>212.0	>212.0	08/06/2021	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115731

WorkList Name : fp8621 WorkList ID : 151224 Department : Wet-Chemistry Date : 08-06-2021 09:00:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/09/2021	Water	M3289-18	Flash Point	Cool 4 deg C	PSEG03	D31	TOTE-COMP	08/04/2021	1010B
08/09/2021	Water	M3293-09	Flash Point	Cool 4 deg C	PSEG03	D31	TOTES-1-2	08/04/2021	1010B
08/09/2021	Water	M3293-11	Flash Point	Cool 4 deg C	PSEG03	D31	2055-57	08/04/2021	1010B
08/09/2021	Water	M3293-13	Flash Point	Cool 4 deg C	PSEG03	D31	2024-29	08/04/2021	1010B
08/10/2021	Water	M3301-01	Flash Point	Cool 4 deg C	PSEG03	D32	MH-P17	08/05/2021	1010B

Date/Time 08/06/21 11:40
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: RJ CFA-106

Date/Time 08/06/21 14:40
 Raw Sample Received by: RJ CFA-106
 Raw Sample Relinquished by: KS(WC)