



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115732

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	08/06/2021	09:50
2	CAL2	1	Water	NA	NA	21.6	7.01	08/06/2021	09:51
3	CAL3	1	Water	NA	NA	21.6	10.03	08/06/2021	09:52
4	ICV	1	Water	NA	NA	21.6	6.99	08/06/2021	10:00
5	CCV1	1	Water	NA	NA	21.6	2.01	08/06/2021	10:01
6	M3316-02	1	Water	NA	NA	20.6	7.49	08/06/2021	10:02
7	M3316-02DUP	1	Water	NA	NA	20.7	7.50	08/06/2021	10:03
8	CCV2	1	Water	NA	NA	21.7	12.02	08/06/2021	10:05

WORKLIST(Hardcopy Internal Chain)

115732

WorkList Name : CORROSIONITY M3316 W WorkList ID : 151229 Department : Wet-Chemistry Date : 08-06-2021 09:40:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2021	Water	M3316-02	Corrosivity	Cool 4 deg C	SEAC01	L11	GROUND-WATER	08/05/2021	9040C

Date/Time 08/06/21 09:45
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 08/06/21 10:50
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)