


LB115735

SFAM01.1

INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 2.1917e-004 *slope*C= -6.5826e-002 *y-intercept*

Rho= 0.9998842

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>Q/D</i>
0.00	0.000	-0.017	-0.017	225	0.000	225					
0.05	0.050					573					
0.20	0.200	0.198	-0.002	1203	0.0 %	1203					
2.50	2.500	2.589	0.089	12112	0.0 %	12112					
5.00	5.000	4.899	-0.101	22651	0.0 %	22651					
7.50	7.500	7.518	0.018	34603	0.0 %	34603					
10.0	10.000	10.013	0.013	45985	0.0 %	45985					

Q/D

1
4
2
0
0

LB115735
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	225	-		0 SFAM01.1	PPB	8/6/2021 14:10	Std
	0.2 S01	1203	-		0.2 SFAM01.1	PPB	8/6/2021 14:13	Std
	2.5 S02	12112	-		2.5 SFAM01.1	PPB	8/6/2021 14:15	Std
	5 S03	22651	-		5 SFAM01.1	PPB	8/6/2021 14:21	Std
	7.5 S04	34603	-		7.5 SFAM01.1	PPB	8/6/2021 14:26	Std
	10 S05	45985	-		10 SFAM01.1	PPB	8/6/2021 14:31	Std
ICV056	ICV056	18468	3.9819 -		SFAM01.1	PPB	8/6/2021 14:34	SMPL
ICB056	ICB056	126	-0.0382 -		SFAM01.1	PPB	8/6/2021 14:36	SMPL
CCV010	CCV010	23091	4.9951 -		SFAM01.1	PPB	8/6/2021 14:41	SMPL
CCB010	CCB010	113	-0.0411 -		SFAM01.1	PPB	8/6/2021 14:43	SMPL
PB138202BL	PBS202	66	-0.0514 -		SFAM01.1	PPB	8/6/2021 14:45	SMPL
M3212-01	MH0AA0	2290	0.4361 -		SFAM01.1	PPB	8/6/2021 14:47	SMPL
M3212-02	MH0AA1	598497	HIGH -		SFAM01.1	PPB	8/6/2021 14:50	
M3212-03	MH0AA2	295895	64.7868 -		SFAM01.1	PPB	8/6/2021 14:52	SMPL
M3212-04	MH0AA3	77149	16.8433 -		SFAM01.1	PPB	8/6/2021 14:54	SMPL
M3212-05	MH0AA4	56163	12.2437 -		SFAM01.1	PPB	8/6/2021 14:56	SMPL
M3212-06	MH0AA5	70430	15.3706 -		SFAM01.1	PPB	8/6/2021 14:59	SMPL
M3212-07	MH0AA6	95	-0.045 -		SFAM01.1	PPB	8/6/2021 15:01	SMPL
M3212-08	MH0AA7	316	0.0034 -		SFAM01.1	PPB	8/6/2021 15:03	SMPL
M3212-09	MH0AA8	109975	24.0379 -		SFAM01.1	PPB	8/6/2021 15:05	SMPL
M3212-10	MH0AB2	71049	15.5063 -		SFAM01.1	PPB	8/6/2021 15:08	SMPL
M3212-11	MH0AB2D	71656	15.6393 -		SFAM01.1	PPB	8/6/2021 15:10	SMPL
M3212-12	MH0AB2S	78752	17.1946 -		SFAM01.1	PPB	8/6/2021 15:15	SMPL
PB138203BL	PBS203	128	-0.0378 -		SFAM01.1	PPB	8/6/2021 15:20	SMPL
M3213-01DLX10	MH0AC5	19988	4.315 -		SFAM01.1	PPB	8/6/2021 15:22	SMPL
M3213-02DLX10	MH0AC6	16286	3.5036 -		SFAM01.1	PPB	8/6/2021 15:24	SMPL
M3213-03DLX10	MH0AC8	7590	1.5977 -		SFAM01.1	PPB	8/6/2021 15:26	SMPL
M3213-04DLX10	MH0AD2	-30	-0.0724 -		SFAM01.1	PPB	8/6/2021 15:29	SMPL
M3213-05DLX10	MH0AD3	23744	5.1382 -		SFAM01.1	PPB	8/6/2021 15:31	SMPL
M3213-06DLX10	MH0AD4	15258	3.2783 -		SFAM01.1	PPB	8/6/2021 15:33	SMPL
CCV011	CCV011	22756	4.9217 -		SFAM01.1	PPB	8/6/2021 15:36	SMPL
CCB011	CCB011	122	-0.0391 -		SFAM01.1	PPB	8/6/2021 15:38	SMPL
M3213-07DLX10	MH0AD5	25580	5.5407 -		SFAM01.1	PPB	8/6/2021 15:42	SMPL
M3213-08DLX10	MH0AD6	406374	HIGH -		SFAM01.1	PPB	8/6/2021 15:44	
M3213-09DLX10	MH0AD7	427961	HIGH -		SFAM01.1	PPB	8/6/2021 15:46	
M3213-10DLX10	MH0AD8	321268	70.3479 -		SFAM01.1	PPB	8/6/2021 15:48	SMPL
M3213-11DLX10	MH0AD9	41194	8.9628 -		SFAM01.1	PPB	8/6/2021 15:50	SMPL
M3213-12DLX10	MH0AE0	88520	19.3355 -		SFAM01.1	PPB	8/6/2021 15:53	SMPL
M3213-13DLX10	MH0AE1	29218	6.338 -		SFAM01.1	PPB	8/6/2021 15:55	SMPL
M3213-14DLX10	MH0AE2	10177	2.1647 -		SFAM01.1	PPB	8/6/2021 15:57	SMPL
M3213-15DLX10	MH0AE3	993485	HIGH -		SFAM01.1	PPB	8/6/2021 15:59	
M3213-16DLX10	MH0AE4	296051	HIGH -		SFAM01.1	PPB	8/6/2021 16:01	
M3213-17DLX10	MH0AE5	703701	HIGH -		SFAM01.1	PPB	8/6/2021 16:07	
CCV012	CCV012	22068	4.7709 -		SFAM01.1	PPB	8/6/2021 16:09	SMPL
CCB012	CCB012	354	0.0118 -		SFAM01.1	PPB	8/6/2021 16:12	SMPL
M3213-18DLX10	MH0AE6	673289	HIGH -		SFAM01.1	PPB	8/6/2021 16:14	

LB115735
INSTRUMENT ID : CV1

M3213-19DLX10	MH0AE7	488084	HIGH	-	SFAM01.1	PPB	8/6/2021 16:16
M3213-20DLX10	MH0AE7D	460330	HIGH	-	SFAM01.1	PPB	8/6/2021 16:20
M3213-21DLX10	MH0AE7S	474777	HIGH	-	SFAM01.1	PPB	8/6/2021 16:22
M3212-02DLX10	MH0AA1	118153	25.8303	-	SFAM01.1	PPB	8/6/2021 16:24 SMPL
M3212-03DLX10	MH0AA2	31731	6.8888	-	SFAM01.1	PPB	8/6/2021 16:26 SMPL
M3212-04DLX5	MH0AA3	15318	3.2915	-	SFAM01.1	PPB	8/6/2021 16:29 SMPL
M3212-05DLX5	MH0AA4	10917	2.3269	-	SFAM01.1	PPB	8/6/2021 16:31 SMPL
M3212-06DLX5	MH0AA5	13619	2.9191	-	SFAM01.1	PPB	8/6/2021 16:33 SMPL
M3212-09DLX5	MH0AA8	21843	4.7216	-	SFAM01.1	PPB	8/6/2021 16:36 SMPL
M3212-10DLX5	MH0AB2	13931	2.9875	-	SFAM01.1	PPB	8/6/2021 16:38 SMPL
M3212-11DLX5	MH0AB2D	14264	3.0605	-	SFAM01.1	PPB	8/6/2021 16:40 SMPL
M3212-12DLX5	MH0AB2S	16556	3.5628	-	SFAM01.1	PPB	8/6/2021 16:42 SMPL
CCV013	CCV013	23175	5.0135	-	SFAM01.1	PPB	8/6/2021 16:45 SMPL
CCB013	CCB013	28	-0.0597	-	SFAM01.1	PPB	8/6/2021 16:47 SMPL
M3213-10DL2X100	MH0AD8	33023	7.172	-	SFAM01.1	PPB	8/6/2021 16:54 SMPL
M3213-12DL2X50	MH0AE0	14788	3.1753	-	SFAM01.1	PPB	8/6/2021 16:56 SMPL
M3213-16DL2X100	MH0AE4	40200	8.745	-	SFAM01.1	PPB	8/6/2021 17:00 SMPL
CCV014	CCV014	22625	4.893	-	SFAM01.1	PPB	8/6/2021 17:14 SMPL
CCB014	CCB014	-92	-0.086	-	SFAM01.1	PPB	8/6/2021 17:16 SMPL
M3213-08DL2X500		11174	2.3832	-	SFAM01.1	PPB	8/6/2021 17:21 SMPL
M3213-09DL2X500		12103	2.5868	-	SFAM01.1	PPB	8/6/2021 17:23 SMPL
M3213-15DL2X500		81199	17.7309	-	SFAM01.1	PPB	8/6/2021 17:26 SMPL
M3213-17DL2X500		27346	5.9277	-	SFAM01.1	PPB	8/6/2021 17:28 SMPL
M3213-18DL2X500		24980	5.4091	-	SFAM01.1	PPB	8/6/2021 17:30 SMPL
M3213-19DL2X500		14696	3.1552	-	SFAM01.1	PPB	8/6/2021 17:32 SMPL
M3213-20DL2X500		14763	3.1698	-	SFAM01.1	PPB	8/6/2021 17:35 SMPL
M3213-21DL2X500		14281	3.0642	-	SFAM01.1	PPB	8/6/2021 17:41 SMPL
M3212-02DLX2X50		23359	5.0539	-	SFAM01.1	PPB	8/6/2021 17:50 SMPL
M3213-15DL2X2500		15661	3.3667	-	SFAM01.1	PPB	8/6/2021 17:57 SMPL
CCV015	CCV015	23122	5.0019	-	SFAM01.1	PPB	8/6/2021 17:59 SMPL
CCB015	CCB015	-52	-0.0772	-	SFAM01.1	PPB	8/6/2021 18:01 SMPL