



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115738

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	08/06/2021	12:05
2	M3293-05		1	9	100.00	>212.0	>212.0	08/06/2021	15:05
3	M3293-05DUP		1	9	100.00	>212.0	>212.0	08/06/2021	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115738

WorkList Name : fp8621-1 WorkList ID : 151225 Department : Wet-Chemistry Date : 08-06-2021 09:01:37

Due Date	Matrix	Sample	Test	Flash Point	Cool 4 deg C	PSEG03	D31	Raw Sample Storage Location	Customer	Customer Sample	Collect Date	Method
08/09/2021	Solid	M3293-05	✓	Flash Point	Cool 4 deg C	PSEG03	D31			1927	08/04/2021	1010A

Date/Time 08/06/21 14:30
Raw Sample Received by: KS (w.c)
Raw Sample Relinquished by: RSCFA-Lab

Date/Time 08/06/21 15:45
Raw Sample Received by: RSCFA-Lab
Raw Sample Relinquished by: KS (w.c)