



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115741

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/06/2021	16:10
2	CAL2	1	Water	NA	NA	21.3	7.00	08/06/2021	16:11
3	CAL3	1	Water	NA	NA	21.3	10.03	08/06/2021	16:12
4	ICV	1	Water	NA	NA	21.4	7.02	08/06/2021	16:15
5	CCV1	1	Water	NA	NA	21.3	2.01	08/06/2021	16:17
6	M3326-01	1	Solid	20.03	20	24.4	9.91	08/06/2021	16:19
7	M3326-01DUP	1	Solid	20.04	20	24.5	9.92	08/06/2021	16:20
8	CCV2	1	Water	NA	NA	21.4	12.02	08/06/2021	16:22

VB 11574)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3326 WorkList ID : 151250 Department : Wet-Chemistry Date : 08-06-2021 15:58:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2021	Solid	M3326-01	pH	Cool 4 deg C	VERI01	L11	MDE-NB	08/05/2021	9045D

Date/Time 08/06/21 15:59
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 08/06/21 17:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)