



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115742

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/06/2021	16:10
2	CAL2	1	Water	NA	NA	21.3	7.00	08/06/2021	16:11
3	CAL3	1	Water	NA	NA	21.3	10.03	08/06/2021	16:12
4	ICV	1	Water	NA	NA	21.4	7.02	08/06/2021	16:15
5	CCV1	1	Water	NA	NA	21.3	2.01	08/06/2021	16:17
6	M3316-01	1	Solid	20.02	20	23.9	6.88	08/06/2021	16:25
7	M3316-01DUP	1	Solid	20.01	20	23.8	6.89	08/06/2021	16:26
8	CCV2	1	Water	NA	NA	21.4	12.02	08/06/2021	16:30

JD 115742

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3233 s WorkList ID : 151251 Department : Wet-Chemistry Date : 08-06-2021 15:59:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2021	Solid	M3316-01	Corrosivity	Cool 4 deg C	SEAC01	L11	SOIL	08/05/2021	9045D

Date/Time 08/06/21 11:00
Raw Sample Received by: SP (WOC)
Raw Sample Relinquished by: JD (SM)

Date/Time 08/06/21 17:15
Raw Sample Received by: JD (SM)
Raw Sample Relinquished by: JP (WOC)