


LB115749

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0800e-004 *slope*C= -4.3659e-002 *y-intercept*

Rho= 0.9999556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.027	-0.027	80	0.000	80					=
0.05	0.050					573					11
0.20	0.200	0.221	0.021	1274	0.0 %	1274					2
2.50	2.500	2.548	0.048	12462	0.0 %	12462					-1
5.00	5.000	4.955	-0.045	24033	0.0 %	24033					0
7.50	7.500	7.472	-0.028	36132	0.0 %	36132					0
10.0	10.000	10.030	0.030	48433	0.0 %	48433					

LB115749

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	80	-		0 SFAM01.1	PPB	8/7/2021 12:21	Std
	0.2 S01	1274	-		0.2 SFAM01.1	PPB	8/7/2021 12:24	Std
	2.5 S02	12462	-		2.5 SFAM01.1	PPB	8/7/2021 12:26	Std
	5 S03	24033	-		5 SFAM01.1	PPB	8/7/2021 12:28	Std
	7.5 S04	36132	-		7.5 SFAM01.1	PPB	8/7/2021 12:30	Std
	10 S05	48433	-		10 SFAM01.1	PPB	8/7/2021 12:33	Std
ICV057	ICV057	19315	3.9738 -		SFAM01.1	PPB	8/7/2021 12:35	SMPL
ICB057	ICB057	-47	-0.0534 -		SFAM01.1	PPB	8/7/2021 12:38	SMPL
CCV016	CCV016	24191	4.988 -		SFAM01.1	PPB	8/7/2021 12:40	SMPL
CCB016	CCB016	-54	-0.0549 -		SFAM01.1	PPB	8/7/2021 12:42	SMPL
M3212-05DLX2	MH0AA4	30198	6.2375 -		SFAM01.1	PPB	8/7/2021 12:44	SMPL
M3213-08DL2X250	MH0AD6	23124	4.7661 -		SFAM01.1	PPB	8/7/2021 12:53	SMPL
CCV017	CCV017	24122	4.9737 -		SFAM01.1	PPB	8/7/2021 12:55	SMPL
CCB017	CCB017	-66	-0.0574 -		SFAM01.1	PPB	8/7/2021 12:57	SMPL